

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092021\
 Data File : PP039392.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Sep 2021 20:33
 Operator : AJ\MA
 Sample : M3792-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 RT4213

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 03:59:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 20 01:59:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.047	4.123	370408	389555	14.768	14.364
2) SA Decachlor...	11.167	9.495	439318	468429	21.244	17.353

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092021\
Data File : PP039392.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Sep 2021 20:33
Operator : AJ\MA
Sample : M3792-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
RT4213

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 03:59:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091721.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 20 01:59:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.047 min
Delta R.T.: 0.000 min
Response: 370408
Conc: 14.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
RT4213




#1 Tetrachloro-m-xylene

R.T.: 4.123 min
Delta R.T.: 0.002 min
Response: 389555
Conc: 14.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 11.167 min
Delta R.T.: -0.005 min
Response: 439318
Conc: 21.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.495 min
Delta R.T.: -0.005 min
Response: 468429
Conc: 17.35 ng/ml