

Data Path : P:\HPCHEM1\ECD_P\Data\PP092115\
 Data File : PP011784.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Sep 2015 18:46
 Operator : IZ/UA
 Sample : G3748-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MGP205-10-12

Manual Integrations
 APPROVED

ugo
 9/22/2015 11:59:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 04:08:04 2015
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP090315.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 19 00:23:08 2015
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.134	4.239	20445773	4596631	467.573m	80.050m#
2) SA Decachlor...	9.512f	9.275	347273	338794	9.079m	5.965m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_P\Data\PP092115\
Data File : PP011784.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Sep 2015 18:46
Operator : IZ/UA
Sample : G3748-04
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
MGP205-10-12


Manual Integrations
APPROVED

ugo
9/22/2015 11:59:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 04:08:04 2015
Quant Method : P:\HPCHEM1\ECD_P\methods\PP090315.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 19 00:23:08 2015
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.134 min
Delta R.T.: -0.037 min
Response: 20445773
Conc: 467.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
MGP205-10-12


Manual Integrations
APPROVED

ugo
9/22/2015 11:59:14 AM




#1 Tetrachloro-m-xylene

R.T.: 4.239 min
Delta R.T.: 0.049 min
Response: 4596631
Conc: 80.05 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.512 min
Delta R.T.: -1.989 min
Response: 347273
Conc: 9.08 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.275 min
Delta R.T.: -0.014 min
Response: 338794
Conc: 5.96 ng/ml m