

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092121\
 Data File : PP039413.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Sep 2021 11:52
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 05:49:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 20 01:59:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.049	4.120	1322169	1391770	52.716	51.320
2) SA Decachlor...	11.172	9.494	1067240	1327258	51.609	49.167
Target Compounds						
21) L5 AR-1248-1	6.374	5.392	274030	292884	543.897	533.333
22) L5 AR-1248-2	6.669	5.658	370985	365491	520.514	518.300
23) L5 AR-1248-3	6.888	5.703	404669	383049	510.647	519.903
24) L5 AR-1248-4	7.318	5.890	434467	447376	502.614	520.995
25) L5 AR-1248-5	7.357	6.316	418508	471829	505.732	525.044



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092121\
Data File : PP039413.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Sep 2021 11:52
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 05:49:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091721.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 20 01:59:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.049 min
Delta R.T.: 0.002 min
Response: 1322169
Conc: 52.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.120 min
Delta R.T.: -0.001 min
Response: 1391770
Conc: 51.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.172 min
Delta R.T.: 0.000 min
Response: 1067240
Conc: 51.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.494 min
Delta R.T.: -0.006 min
Response: 1327258
Conc: 49.17 ng/ml



#21 AR-1248-1


R.T.: 6.374 min
Delta R.T.: 0.000 min
Response: 274030
Conc: 543.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.392 min
Delta R.T.: -0.003 min
Response: 292884
Conc: 533.33 ng/ml




#22 AR-1248-2

R.T.: 6.669 min
Delta R.T.: 0.000 min
Response: 370985
Conc: 520.51 ng/ml



#22 AR-1248-2


R.T.: 5.658 min
Delta R.T.: -0.004 min
Response: 365491
Conc: 518.30 ng/ml



#23 AR-1248-3


R.T.: 6.888 min
Delta R.T.: 0.000 min
Response: 404669
Conc: 510.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.703 min
Delta R.T.: -0.003 min
Response: 383049
Conc: 519.90 ng/ml




#24 AR-1248-4

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 434467
Conc: 502.61 ng/ml



#24 AR-1248-4


R.T.: 5.890 min
Delta R.T.: -0.003 min
Response: 447376
Conc: 521.00 ng/ml



#25 AR-1248-5

R.T.: 7.357 min
Delta R.T.: 0.000 min
Response: 418508
Conc: 505.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.316 min
Delta R.T.: -0.003 min
Response: 471829
Conc: 525.04 ng/ml