

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092122\
 Data File : PP051514.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2022 00:04
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 01:15:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.682	86449038	76719824	48.647	50.790
2) SA Decachlor...	10.250	8.782	72656636	60406089	47.013	50.414
Target Compounds						
26) L6 AR-1254-1	6.462	5.602	29024036	39201020	467.151	482.595
27) L6 AR-1254-2	6.682	5.751	51884807	33584515	449.485	477.888
28) L6 AR-1254-3	7.050	6.161	57948204	54041774	476.024	481.537
29) L6 AR-1254-4	7.335	6.392	39847370	31329068	481.147	482.665
30) L6 AR-1254-5	7.757	6.815	43464121	45107827	487.847	482.450



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092122\
Data File : PP051514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2022 00:04
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 01:15:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: -0.004 min
Response: 86449038
Conc: 48.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.002 min
Response: 76719824
Conc: 50.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: -0.004 min
Response: 72656636
Conc: 47.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.782 min
Delta R.T.: -0.005 min
Response: 60406089
Conc: 50.41 ng/ml



#26 AR-1254-1


R.T.: 6.462 min
Delta R.T.: -0.005 min
Response: 29024036
Conc: 467.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.602 min
Delta R.T.: -0.005 min
Response: 39201020
Conc: 482.60 ng/ml




#27 AR-1254-2

R.T.: 6.682 min
Delta R.T.: -0.005 min
Response: 51884807
Conc: 449.49 ng/ml



#27 AR-1254-2


R.T.: 5.751 min
Delta R.T.: -0.005 min
Response: 33584515
Conc: 477.89 ng/ml



#28 AR-1254-3


R.T.: 7.050 min
Delta R.T.: -0.005 min
Response: 57948204
Conc: 476.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.161 min
Delta R.T.: -0.006 min
Response: 54041774
Conc: 481.54 ng/ml




#29 AR-1254-4

R.T.: 7.335 min
Delta R.T.: -0.006 min
Response: 39847370
Conc: 481.15 ng/ml



#29 AR-1254-4


R.T.: 6.392 min
Delta R.T.: -0.005 min
Response: 31329068
Conc: 482.66 ng/ml



#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: -0.006 min
Response: 43464121
Conc: 487.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.815 min
Delta R.T.: -0.006 min
Response: 45107827
Conc: 482.45 ng/ml