

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
 Data File : PP051566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2022 09:25
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/26/2022
 Supervised By :Ankita Jodhani 09/26/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 11:18:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.429	3.680	84826952	74425756	47.734	49.271m
2)	SA Decachlor...	10.250	8.780	63991193	56195253	41.406	46.899
Target Compounds							
3)	L1 AR-1016-1	5.603	4.784	30113304	24135321	488.637	484.274
4)	L1 AR-1016-2	5.625	4.803	42829463	34464651	495.253	482.774
5)	L1 AR-1016-3	5.687	4.982	26832676	18565056	490.405	494.424
6)	L1 AR-1016-4	5.786	5.023	21150332	14870101	496.747	489.236
7)	L1 AR-1016-5	6.083	5.242	22185213	18932026	487.424	477.578
31)	L7 AR-1260-1	7.215	6.291	42968967	35897433	460.278	481.110
32)	L7 AR-1260-2	7.473	6.480	47142824	41695310	446.502	471.572
33)	L7 AR-1260-3	7.833	6.637	36386789	39340335	471.057	463.864
34)	L7 AR-1260-4	8.060	7.114	39821802	31492292	448.743	469.716
35)	L7 AR-1260-5	8.384	7.356	67478300	66412354	438.609	449.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
Data File : PP051566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2022 09:25
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/26/2022
Supervised By :Ankita Jodhani 09/26/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 11:18:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

