

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
 Data File : PP051581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2022 14:19
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/26/2022
 Supervised By :Ankita Jodhani 09/26/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 15:47:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.428	3.680	92374475	79093570	51.982	52.361m
2)	SA Decachlor...	10.252	8.782	65247058	58863900	42.219	49.126
Target Compounds							
3)	L1 AR-1016-1	5.602	4.785	31432626	25451128	510.045	510.675
4)	L1 AR-1016-2	5.625	4.804	45049237	36142460	520.921	506.276
5)	L1 AR-1016-3	5.687	4.984	28323300	19476143	517.649	518.688
6)	L1 AR-1016-4	5.786	5.026	22334482	15541637	524.558	511.330
7)	L1 AR-1016-5	6.084	5.243	23257007	20084290	510.972	506.644
31)	L7 AR-1260-1	7.215	6.292	43709400	37208760	468.210	498.685
32)	L7 AR-1260-2	7.472	6.482	48293398	43450690	457.399	491.425
33)	L7 AR-1260-3	7.834	6.638	36526186	41158640	472.861	485.303
34)	L7 AR-1260-4	8.061	7.115	40233255	32957005	453.380	491.563
35)	L7 AR-1260-5	8.385	7.358	69109398	72298395	449.211	489.356

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
Data File : PP051581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2022 14:19
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/26/2022
Supervised By :Ankita Jodhani 09/26/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:47:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

