

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
 Data File : PP051624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2022 04:35
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 06:42:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.679	93711301	80037052	52.734	52.986
2) SA Decachlor...	10.249	8.776	70151057	63407376	45.392	52.918
Target Compounds						
26) L6 AR-1254-1	6.460	5.598	31247433	41555922	502.938	511.586
27) L6 AR-1254-2	6.680	5.747	51537346	35405996	446.475	503.807
28) L6 AR-1254-3	7.048	6.157	52833628	56044919	434.010	499.385
29) L6 AR-1254-4	7.334	6.388	37862107	33004638	457.176	508.479
30) L6 AR-1254-5	7.755	6.811	41735262	47156893	468.442	504.366



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092322\
Data File : PP051624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2022 04:35
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 06:42:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: -0.006 min
Response: 93711301
Conc: 52.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.679 min
Delta R.T.: -0.006 min
Response: 80037052
Conc: 52.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.249 min
Delta R.T.: -0.005 min
Response: 70151057
Conc: 45.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.776 min
Delta R.T.: -0.011 min
Response: 63407376
Conc: 52.92 ng/ml



#26 AR-1254-1


R.T.: 6.460 min
Delta R.T.: -0.008 min
Response: 31247433
Conc: 502.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.598 min
Delta R.T.: -0.009 min
Response: 41555922
Conc: 511.59 ng/ml




#27 AR-1254-2

R.T.: 6.680 min
Delta R.T.: -0.008 min
Response: 51537346
Conc: 446.48 ng/ml



#27 AR-1254-2


R.T.: 5.747 min
Delta R.T.: -0.009 min
Response: 35405996
Conc: 503.81 ng/ml



#28 AR-1254-3


R.T.: 7.048 min
Delta R.T.: -0.006 min
Response: 52833628
Conc: 434.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.157 min
Delta R.T.: -0.010 min
Response: 56044919
Conc: 499.39 ng/ml




#29 AR-1254-4

R.T.: 7.334 min
Delta R.T.: -0.007 min
Response: 37862107
Conc: 457.18 ng/ml



#29 AR-1254-4


R.T.: 6.388 min
Delta R.T.: -0.010 min
Response: 33004638
Conc: 508.48 ng/ml



#30 AR-1254-5

R.T.: 7.755 min
Delta R.T.: -0.007 min
Response: 41735262
Conc: 468.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.811 min
Delta R.T.: -0.010 min
Response: 47156893
Conc: 504.37 ng/ml