

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP092418.M
 Title : 8080.M
 Last Update : Tue Sep 25 01:51:01 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP027481.D 1000 =PP027482.D 1500 =PP027483.D
 500 =PP027480.D 200 =PP027479.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.062	5.897	5.745	5.976	6.184	5.973	E5	2.78
2) T	3,5-DICHLOROBENZOI...	5.031	4.791	4.648	4.993	5.264	4.946	E5	4.78
3) T	4-Nitrophenol	4.555	4.391	4.415	4.324	4.239	4.385	E5	2.67
4) S	2,4-DCAA	2.606	2.489	2.438	2.572	2.575	2.536	E5	2.74
5) T	DICAMBA	1.577	1.510	1.465	1.559	1.628	1.548	E6	4.05
6) T	MCPP	4.668	4.642	4.811	3.957	2.806	4.177	E5	19.99
7) T	MCPA	7.335	7.467	7.529	6.420	4.514	6.653	E5	19.20
8) T	DICHLORPROP	3.727	3.545	3.395	3.720	3.730	3.623	E5	4.13
9) T	2,4-D	3.994	3.847	3.781	3.881	3.847	3.870	E5	2.02
10) T	Pentachlorophenol	8.234	7.710	7.129	8.400	8.890	8.073	E6	8.37
11) T	2,4,5-TP (SILVEX)	2.986	2.834	2.702	2.973	3.100	2.919	E6	5.26
12) T	2,4,5-T	2.577	2.473	2.390	2.519	2.537	2.499	E6	2.87
13) T	2,4-DB	2.009	1.996	2.001	1.852	1.618	1.895	E5	8.87
14) T	DINOSEB	1.982	1.889	1.863	1.920	1.971	1.925	E6	2.66
15) T	Picloram	3.281	3.279	3.395	2.888	2.522	3.073	E6	11.82
16) T	DCPA	4.430	4.198	3.973	4.443	4.651	4.339	E6	5.99

Signal #2 Calibration Files

750 =PP027481.D 1000 =PP027482.D 1500 =PP027483.D
 500 =PP027480.D 200 =PP027479.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.544	5.353	5.256	5.435	5.607	5.439	E5	2.60
2) T	3,5-DICHLOROBENZOI...	5.370	5.152	5.030	5.324	5.546	5.284	E5	3.78
3) T	4-Nitrophenol	2.129	2.068	2.086	1.972	1.800	2.011	E5	6.53
4) S	2,4-DCAA	4.405	4.228	4.130	4.342	4.538	4.328	E5	3.64
5) T	DICAMBA	1.689	1.613	1.549	1.672	1.746	1.654	E6	4.55
6) T	MCPP	8.903	8.818	8.835	8.411	7.173	8.428	E5	8.63
7) T	MCPA	1.423	1.392	1.409	1.338	1.150	1.343	E6	8.35
8) T	DICHLORPROP	4.205	4.002	3.885	4.189	4.463	4.149	E5	5.32
9) T	2,4-D	5.404	5.176	5.063	5.278	5.315	5.247	E5	2.50
10) T	Pentachlorophenol	6.688	6.196	5.663	6.910	7.519	6.595	E6	10.69
11) T	2,4,5-TP (SILVEX)	2.575	2.436	2.312	2.588	2.713	2.525	E6	6.10
12) T	2,4,5-T	2.069	1.992	1.941	1.979	1.915	1.979	E6	2.97
13) T	2,4-DB	2.540	2.505	2.543	2.363	2.210	2.432	E5	5.93
14) T	DINOSEB	1.640	1.564	1.542	1.586	1.616	1.590	E6	2.47
15) T	Picloram	2.705	2.676	2.740	2.379	2.017	2.503	E6	12.29
16) T	DCPA	3.244	3.041	2.848	3.291	3.537	3.192	E6	8.18

(#) = Out of Range