

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092421\
 Data File : PP039521.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Sep 2021 17:59
 Operator : AJ\MA
 Sample : PP092421ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092421

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 18:33:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 18:31:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.906	3.996	1504047	783962	47.510	48.113
2)	SA Decachlor...	10.930	9.336	1024643	1043270	47.232	47.742
Target Compounds							
3)	L1 AR-1016-1	6.229	5.263	511212	336097	476.797	466.359
4)	L1 AR-1016-2	6.252	5.283	731884	461846	466.563	468.084
5)	L1 AR-1016-3	6.318	5.476	461020	256288	470.821	470.094
6)	L1 AR-1016-4	6.426	5.528	379748	207342	484.998	467.947
7)	L1 AR-1016-5	6.742	5.759	362842	264889	490.227	479.141
8)	L2 AR-1221-1	5.149	4.257	48999	28691	130.558	142.404
9)	L2 AR-1221-2	5.256	4.357	55361	42108	203.358	247.092
10)	L2 AR-1221-3	5.342	4.446	283409	179030	331.070	331.122
11)	L3 AR-1232-1	5.342	4.446	283409	179030	442.830	442.713
12)	L3 AR-1232-2	5.932	5.283	389778	461846	1082.742	1105.970
13)	L3 AR-1232-3	6.252	5.476	731884	256288	1124.798	1155.223
14)	L3 AR-1232-4	6.426	5.573	379748	259234	1132.907	1104.530
15)	L3 AR-1232-5	6.524	5.759	293269	264889	1272.476	1212.580
16)	L4 AR-1242-1	6.229	5.263	511212	336097	631.666	626.612
17)	L4 AR-1242-2	6.252	5.283	731884	461846	627.440	630.596
18)	L4 AR-1242-3	6.318	5.476	461020	256288	629.006	622.405
19)	L4 AR-1242-4	6.426	5.573	379748	259234	634.948	613.658
20)	L4 AR-1242-5	7.210	6.140	54506	209376	87.760	394.564 #
21)	L5 AR-1248-1	6.229	5.263	511212	336097	796.095	789.976
22)	L5 AR-1248-2	6.524	5.528	293269	207342	355.405	349.888
23)	L5 AR-1248-3	6.742	5.573	362842	259234	368.902	427.836
24)	L5 AR-1248-4	7.171	5.759	62696	264889	57.754	389.008 #
25)	L5 AR-1248-5	7.210	6.183	54506	34872	52.233	49.472
26)	L6 AR-1254-1	7.140	6.140	262541	209376	240.530	201.105
27)	L6 AR-1254-2	7.369	6.300	271603	199939	168.148	207.469
28)	L6 AR-1254-3	7.752	6.742f	150885	363924	90.684	241.914 #
29)	L6 AR-1254-4	8.047	6.963	98446	47717	80.592	49.554 #
30)	L6 AR-1254-5	8.478	7.395	778080	685976	613.240	487.360
31)	L7 AR-1260-1	7.920	6.862	591383	503774	481.241	467.795
32)	L7 AR-1260-2	8.186	7.058	672564	594806	485.805	464.403
33)	L7 AR-1260-3	8.551	7.215	414735	559115	501.424	464.752
34)	L7 AR-1260-4	8.782	7.701	483520	448639	484.992	481.717
35)	L7 AR-1260-5	9.107	7.948	921276	1024674	486.166	477.719
36)	L8 AR-1262-1	8.551	7.395	414735	685976	310.599	973.034 #
37)	L8 AR-1262-2	9.107	7.948	921276	1024674	418.846	451.053
38)	L8 AR-1262-3	9.439f	8.237	592040	265928	606.480	261.641 #
39)	L8 AR-1262-4	9.492f	8.301	342852	774561	573.186	431.223
40)	L8 AR-1262-5	10.178	8.800	219687	290282	282.177	318.832
41)	L9 AR-1268-1	9.439f	8.237	592040	265928	218.179	91.199 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092421\
 Data File : PP039521.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Sep 2021 17:59
 Operator : AJ\MA
 Sample : PP092421ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092421

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 18:33:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 18:31:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.492f	8.301	342852	774561	133.103	290.565 #
43) L9	AR-1268-3	0.000	8.512	0	16842	N.D.	7.340 #
44) L9	AR-1268-4	10.178	8.800	219687	290282	248.265	282.021
45) L9	AR-1268-5	10.592	9.086	85680	101337	10.948	13.810 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092421\
Data File : PP039521.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Sep 2021 17:59
Operator : AJ\MA
Sample : PP092421ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092421

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 18:33:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 18:31:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 1504047
Conc: 47.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#1 Tetrachloro-m-xylene

R.T.: 3.996 min
Delta R.T.: 0.000 min
Response: 783962
Conc: 48.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.930 min
Delta R.T.: 0.000 min
Response: 1024643
Conc: 47.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.336 min
Delta R.T.: 0.000 min
Response: 1043270
Conc: 47.74 ng/ml



#3 AR-1016-1


R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 511212
Conc: 476.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#3 AR-1016-1

R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 336097
Conc: 466.36 ng/ml




#4 AR-1016-2

R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 731884
Conc: 466.56 ng/ml



#4 AR-1016-2


R.T.: 5.283 min
Delta R.T.: 0.000 min
Response: 461846
Conc: 468.08 ng/ml



#5 AR-1016-3


R.T.: 6.318 min
Delta R.T.: 0.001 min
Response: 461020
Conc: 470.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#5 AR-1016-3

R.T.: 5.476 min
Delta R.T.: 0.000 min
Response: 256288
Conc: 470.09 ng/ml




#6 AR-1016-4

R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 379748
Conc: 485.00 ng/ml



#6 AR-1016-4


R.T.: 5.528 min
Delta R.T.: 0.000 min
Response: 207342
Conc: 467.95 ng/ml



#7 AR-1016-5


R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 362842
Conc: 490.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#7 AR-1016-5

R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 264889
Conc: 479.14 ng/ml




#8 AR-1221-1

R.T.: 5.149 min
Delta R.T.: -0.007 min
Response: 48999
Conc: 130.56 ng/ml



#8 AR-1221-1


R.T.: 4.257 min
Delta R.T.: -0.002 min
Response: 28691
Conc: 142.40 ng/ml



#9 AR-1221-2


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 55361
Conc: 203.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#9 AR-1221-2

R.T.: 4.357 min
Delta R.T.: -0.001 min
Response: 42108
Conc: 247.09 ng/ml




#10 AR-1221-3

R.T.: 5.342 min
Delta R.T.: 0.000 min
Response: 283409
Conc: 331.07 ng/ml



#10 AR-1221-3


R.T.: 4.446 min
Delta R.T.: -0.001 min
Response: 179030
Conc: 331.12 ng/ml



#11 AR-1232-1


R.T.: 5.342 min
Delta R.T.: 0.000 min
Response: 283409
Conc: 442.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#11 AR-1232-1

R.T.: 4.446 min
Delta R.T.: 0.000 min
Response: 179030
Conc: 442.71 ng/ml




#12 AR-1232-2

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 389778
Conc: 1082.74 ng/ml



#12 AR-1232-2


R.T.: 5.283 min
Delta R.T.: 0.000 min
Response: 461846
Conc: 1105.97 ng/ml



#13 AR-1232-3


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 731884
Conc: 1124.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#13 AR-1232-3

R.T.: 5.476 min
Delta R.T.: 0.000 min
Response: 256288
Conc: 1155.22 ng/ml




#14 AR-1232-4

R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 379748
Conc: 1132.91 ng/ml



#14 AR-1232-4


R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 259234
Conc: 1104.53 ng/ml



#15 AR-1232-5


R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 293269
Conc: 1272.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#15 AR-1232-5

R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 264889
Conc: 1212.58 ng/ml




#16 AR-1242-1

R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 511212
Conc: 631.67 ng/ml



#16 AR-1242-1


R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 336097
Conc: 626.61 ng/ml



#17 AR-1242-2


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 731884
Conc: 627.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#17 AR-1242-2

R.T.: 5.283 min
Delta R.T.: 0.000 min
Response: 461846
Conc: 630.60 ng/ml




#18 AR-1242-3

R.T.: 6.318 min
Delta R.T.: 0.000 min
Response: 461020
Conc: 629.01 ng/ml



#18 AR-1242-3


R.T.: 5.476 min
Delta R.T.: 0.000 min
Response: 256288
Conc: 622.40 ng/ml



#19 AR-1242-4


R.T.: 6.426 min
Delta R.T.: 0.000 min
Response: 379748
Conc: 634.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#19 AR-1242-4

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 259234
Conc: 613.66 ng/ml




#20 AR-1242-5

R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 54506
Conc: 87.76 ng/ml



#20 AR-1242-5


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 209376
Conc: 394.56 ng/ml



#21 AR-1248-1


R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 511212
Conc: 796.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#21 AR-1248-1

R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 336097
Conc: 789.98 ng/ml




#22 AR-1248-2

R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 293269
Conc: 355.41 ng/ml



#22 AR-1248-2


R.T.: 5.528 min
Delta R.T.: 0.000 min
Response: 207342
Conc: 349.89 ng/ml



#23 AR-1248-3


R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 362842
Conc: 368.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#23 AR-1248-3

R.T.: 5.573 min
Delta R.T.: 0.000 min
Response: 259234
Conc: 427.84 ng/ml




#24 AR-1248-4

R.T.: 7.171 min
Delta R.T.: 0.000 min
Response: 62696
Conc: 57.75 ng/ml



#24 AR-1248-4


R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 264889
Conc: 389.01 ng/ml



#25 AR-1248-5


R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 54506
Conc: 52.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#25 AR-1248-5

R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 34872
Conc: 49.47 ng/ml




#26 AR-1254-1

R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 262541
Conc: 240.53 ng/ml



#26 AR-1254-1


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 209376
Conc: 201.10 ng/ml



#27 AR-1254-2


R.T.: 7.369 min
Delta R.T.: -0.002 min
Response: 271603
Conc: 168.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#27 AR-1254-2

R.T.: 6.300 min
Delta R.T.: 0.000 min
Response: 199939
Conc: 207.47 ng/ml




#28 AR-1254-3

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 150885
Conc: 90.68 ng/ml



#28 AR-1254-3


R.T.: 6.742 min
Delta R.T.: 0.017 min
Response: 363924
Conc: 241.91 ng/ml



#29 AR-1254-4


R.T.: 8.047 min
Delta R.T.: -0.002 min
Response: 98446
Conc: 80.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#29 AR-1254-4

R.T.: 6.963 min
Delta R.T.: 0.000 min
Response: 47717
Conc: 49.55 ng/ml




#30 AR-1254-5

R.T.: 8.478 min
Delta R.T.: 0.000 min
Response: 778080
Conc: 613.24 ng/ml



#30 AR-1254-5


R.T.: 7.395 min
Delta R.T.: -0.001 min
Response: 685976
Conc: 487.36 ng/ml



#31 AR-1260-1


R.T.: 7.920 min
Delta R.T.: 0.000 min
Response: 591383
Conc: 481.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#31 AR-1260-1

R.T.: 6.862 min
Delta R.T.: 0.000 min
Response: 503774
Conc: 467.80 ng/ml




#32 AR-1260-2

R.T.: 8.186 min
Delta R.T.: 0.000 min
Response: 672564
Conc: 485.81 ng/ml



#32 AR-1260-2


R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 594806
Conc: 464.40 ng/ml



#33 AR-1260-3


R.T.: 8.551 min
Delta R.T.: 0.002 min
Response: 414735
Conc: 501.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#33 AR-1260-3

R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 559115
Conc: 464.75 ng/ml




#34 AR-1260-4

R.T.: 8.782 min
Delta R.T.: 0.002 min
Response: 483520
Conc: 484.99 ng/ml



#34 AR-1260-4


R.T.: 7.701 min
Delta R.T.: 0.000 min
Response: 448639
Conc: 481.72 ng/ml



#35 AR-1260-5


R.T.: 9.107 min
Delta R.T.: 0.000 min
Response: 921276
Conc: 486.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#35 AR-1260-5

R.T.: 7.948 min
Delta R.T.: 0.000 min
Response: 1024674
Conc: 477.72 ng/ml




#36 AR-1262-1

R.T.: 8.551 min
Delta R.T.: 0.000 min
Response: 414735
Conc: 310.60 ng/ml



#36 AR-1262-1


R.T.: 7.395 min
Delta R.T.: 0.001 min
Response: 685976
Conc: 973.03 ng/ml



#37 AR-1262-2


R.T.: 9.107 min
Delta R.T.: 0.000 min
Response: 921276
Conc: 418.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#37 AR-1262-2

R.T.: 7.948 min
Delta R.T.: 0.000 min
Response: 1024674
Conc: 451.05 ng/ml




#38 AR-1262-3

R.T.: 9.439 min
Delta R.T.: 0.022 min
Response: 592040
Conc: 606.48 ng/ml



#38 AR-1262-3


R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 265928
Conc: 261.64 ng/ml



#39 AR-1262-4


R.T.: 9.492 min
Delta R.T.: -0.019 min
Response: 342852
Conc: 573.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#39 AR-1262-4

R.T.: 8.301 min
Delta R.T.: -0.001 min
Response: 774561
Conc: 431.22 ng/ml




#40 AR-1262-5

R.T.: 10.178 min
Delta R.T.: -0.002 min
Response: 219687
Conc: 282.18 ng/ml



#40 AR-1262-5


R.T.: 8.800 min
Delta R.T.: -0.001 min
Response: 290282
Conc: 318.83 ng/ml



#41 AR-1268-1


R.T.: 9.439 min
Delta R.T.: 0.022 min
Response: 592040
Conc: 218.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#41 AR-1268-1

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 265928
Conc: 91.20 ng/ml




#42 AR-1268-2

R.T.: 9.492 min
Delta R.T.: -0.021 min
Response: 342852
Conc: 133.10 ng/ml



#42 AR-1268-2


R.T.: 8.301 min
Delta R.T.: -0.002 min
Response: 774561
Conc: 290.57 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.740 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
ICVPP092421




#43 AR-1268-3

R.T.: 8.512 min
Delta R.T.: 0.000 min
Response: 16842
Conc: 7.34 ng/ml




#44 AR-1268-4

R.T.: 10.178 min
Delta R.T.: 0.002 min
Response: 219687
Conc: 248.27 ng/ml



#44 AR-1268-4


R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 290282
Conc: 282.02 ng/ml



#45 AR-1268-5

R.T.: 10.592 min
Delta R.T.: -0.002 min
Response: 85680
Conc: 10.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421



#45 AR-1268-5

R.T.: 9.086 min
Delta R.T.: 0.000 min
Response: 101337
Conc: 13.81 ng/ml