

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092421\
 Data File : PP039524.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Sep 2021 18:49
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092421AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 19:11:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 24 18:31:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.906	3.996	1561129	852481	49.313	52.318
2) SA Decachlor...	10.930	9.336	1095492	1101797	50.498	50.420
Target Compounds						
26) L6 AR-1254-1	7.141	6.140	524066	517207	480.129	496.775
27) L6 AR-1254-2	7.370	6.300	767954	471413	475.436	489.167
28) L6 AR-1254-3	7.752	6.724	804417	747302	483.464	496.761
29) L6 AR-1254-4	8.049	6.963	584997	474210	478.904	492.462
30) L6 AR-1254-5	8.479	7.396	618994	702808	487.857	499.318



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092421\
Data File : PP039524.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Sep 2021 18:49
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092421AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 19:11:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 24 18:31:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 1561129
Conc: 49.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.996 min
Delta R.T.: 0.000 min
Response: 852481
Conc: 52.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.930 min
Delta R.T.: 0.000 min
Response: 1095492
Conc: 50.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.336 min
Delta R.T.: 0.000 min
Response: 1101797
Conc: 50.42 ng/ml



#26 AR-1254-1


R.T.: 7.141 min
Delta R.T.: 0.000 min
Response: 524066
Conc: 480.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421AR1254




#26 AR-1254-1

R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 517207
Conc: 496.77 ng/ml




#27 AR-1254-2

R.T.: 7.370 min
Delta R.T.: 0.000 min
Response: 767954
Conc: 475.44 ng/ml



#27 AR-1254-2


R.T.: 6.300 min
Delta R.T.: -0.001 min
Response: 471413
Conc: 489.17 ng/ml



#28 AR-1254-3


R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 804417
Conc: 483.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421AR1254




#28 AR-1254-3

R.T.: 6.724 min
Delta R.T.: 0.000 min
Response: 747302
Conc: 496.76 ng/ml




#29 AR-1254-4

R.T.: 8.049 min
Delta R.T.: 0.000 min
Response: 584997
Conc: 478.90 ng/ml



#29 AR-1254-4


R.T.: 6.963 min
Delta R.T.: 0.000 min
Response: 474210
Conc: 492.46 ng/ml



#30 AR-1254-5

R.T.: 8.479 min
Delta R.T.: 0.000 min
Response: 618994
Conc: 487.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092421AR1254



#30 AR-1254-5

R.T.: 7.396 min
Delta R.T.: 0.000 min
Response: 702808
Conc: 499.32 ng/ml