

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092525\
 Data File : PP075314.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2025 09:48
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 08:15:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.798	69953867	325.1E6	53.462	53.281
2) SA Decachlor...	10.438	8.803	55583223	395.7E6	55.720	65.596
Target Compounds						
26) L6 AR-1254-1	6.658	5.676	33514355	267.8E6	521.835	536.895
27) L6 AR-1254-2	6.874	5.823	45035768	211.0E6	524.075	528.553
28) L6 AR-1254-3	7.236	6.225	47893737	377.1E6	518.205	567.243
29) L6 AR-1254-4	7.518	6.452	37412138	277.6E6	524.352	562.408
30) L6 AR-1254-5	7.934	6.867	44804263	322.1E6	547.680	564.483



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092525\
Data File : PP075314.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2025 09:48
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 08:15:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.658 min
Delta R.T.: 0.003 min
Response: 69953867
Conc: 53.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.002 min
Response: 325059411
Conc: 53.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.438 min
Delta R.T.: 0.003 min
Response: 55583223
Conc: 55.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 395706317
Conc: 65.60 ng/ml



#26 AR-1254-1


R.T.: 6.658 min
Delta R.T.: 0.000 min
Response: 33514355
Conc: 521.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.676 min
Delta R.T.: -0.002 min
Response: 267804652
Conc: 536.90 ng/ml




#27 AR-1254-2

R.T.: 6.874 min
Delta R.T.: -0.002 min
Response: 45035768
Conc: 524.08 ng/ml



#27 AR-1254-2


R.T.: 5.823 min
Delta R.T.: -0.002 min
Response: 210989974
Conc: 528.55 ng/ml



#28 AR-1254-3


R.T.: 7.236 min
Delta R.T.: -0.002 min
Response: 47893737
Conc: 518.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.225 min
Delta R.T.: -0.002 min
Response: 377060075
Conc: 567.24 ng/ml




#29 AR-1254-4

R.T.: 7.518 min
Delta R.T.: -0.002 min
Response: 37412138
Conc: 524.35 ng/ml



#29 AR-1254-4


R.T.: 6.452 min
Delta R.T.: -0.001 min
Response: 277631207
Conc: 562.41 ng/ml



#30 AR-1254-5

R.T.: 7.934 min
Delta R.T.: -0.002 min
Response: 44804263
Conc: 547.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.867 min
Delta R.T.: -0.003 min
Response: 322146099
Conc: 564.48 ng/ml