

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092622\
 Data File : PP051629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2022 09:55
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 11:29:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.680	95776357	76442017	53.896	50.606
2) SA Decachlor...	10.253	8.780	68591114	61679853	44.383	51.477
Target Compounds						
21) L5 AR-1248-1	5.602	4.784	20201372	15963366	509.398	493.748
22) L5 AR-1248-2	5.878	5.024	29566328	22046582	521.227	486.067
23) L5 AR-1248-3	6.083	5.068	32618732	23178910	519.515	490.354
24) L5 AR-1248-4	6.488	5.242	28740559	25947108	473.975	464.966
25) L5 AR-1248-5	6.527	5.642	28075326	25776726	449.835	488.960



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092622\
Data File : PP051629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2022 09:55
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 11:29:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: -0.004 min
Response: 95776357
Conc: 53.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: -0.005 min
Response: 76442017
Conc: 50.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.253 min
Delta R.T.: 0.000 min
Response: 68591114
Conc: 44.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.780 min
Delta R.T.: -0.007 min
Response: 61679853
Conc: 51.48 ng/ml



#21 AR-1248-1


R.T.: 5.602 min
Delta R.T.: -0.005 min
Response: 20201372
Conc: 509.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.784 min
Delta R.T.: -0.006 min
Response: 15963366
Conc: 493.75 ng/ml




#22 AR-1248-2

R.T.: 5.878 min
Delta R.T.: -0.004 min
Response: 29566328
Conc: 521.23 ng/ml



#22 AR-1248-2


R.T.: 5.024 min
Delta R.T.: -0.007 min
Response: 22046582
Conc: 486.07 ng/ml



#23 AR-1248-3


R.T.: 6.083 min
Delta R.T.: -0.005 min
Response: 32618732
Conc: 519.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.068 min
Delta R.T.: -0.006 min
Response: 23178910
Conc: 490.35 ng/ml




#24 AR-1248-4

R.T.: 6.488 min
Delta R.T.: -0.005 min
Response: 28740559
Conc: 473.98 ng/ml



#24 AR-1248-4


R.T.: 5.242 min
Delta R.T.: -0.007 min
Response: 25947108
Conc: 464.97 ng/ml



#25 AR-1248-5

R.T.: 6.527 min
Delta R.T.: -0.004 min
Response: 28075326
Conc: 449.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.642 min
Delta R.T.: -0.006 min
Response: 25776726
Conc: 488.96 ng/ml