

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP092622\  
 Data File : PP051646.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 26 Sep 2022 14:54  
 Operator : YP\AJ  
 Sample : PB147889BL  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 PB147889BL

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Sep 26 15:48:04 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP091922.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Tue Sep 20 14:37:22 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|--------|-------|----------|----------|--------|--------|
| -----                       |        |       |          |          |        |        |
| System Monitoring Compounds |        |       |          |          |        |        |
| 1) SA Tetrachlo...          | 4.432  | 3.684 | 36429565 | 29017332 | 20.500 | 19.210 |
| 2) SA Decachlor...          | 10.261 | 8.786 | 29870899 | 26212932 | 19.328 | 21.877 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP092622\  
Data File : PP051646.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 26 Sep 2022 14:54  
Operator : YP\AJ  
Sample : PB147889BL  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
PB147889BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 26 15:48:04 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP091922.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Sep 20 14:37:22 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.432 min  
Delta R.T.: 0.000 min  
Response: 36429565  
Conc: 20.50 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
PB147889BL




#1 Tetrachloro-m-xylene

R.T.: 3.684 min  
Delta R.T.: 0.000 min  
Response: 29017332  
Conc: 19.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.261 min  
Delta R.T.: 0.007 min  
Response: 29870899  
Conc: 19.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.786 min  
Delta R.T.: -0.001 min  
Response: 26212932  
Conc: 21.88 ng/ml