

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092623\
 Data File : PP060252.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2023 17:39
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 18:14:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 26 17:51:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	58026550	38553104	25.822	25.827
2) SA Decachlor...	10.456	8.751	47066267	42803182	27.112	26.324
Target Compounds						
3) L1 AR-1016-1	5.719	4.784	19807608	15556289	272.885	269.199
4) L1 AR-1016-2	5.743	4.803	27770671	21292150	268.147	264.283
5) L1 AR-1016-3	5.805	4.981	17819580	12044638	274.670	268.189
6) L1 AR-1016-4	5.906	5.023	14771441	9787366	272.569	272.859
7) L1 AR-1016-5	6.204	5.239	15303079	13058864	275.891	271.015
31) L7 AR-1260-1	7.343	6.281	28323067	26771388	276.474	268.612
32) L7 AR-1260-2	7.602	6.469	32370783	31426517	276.005	266.163
33) L7 AR-1260-3	7.965	6.625	25220660	30065921	278.284	263.391
34) L7 AR-1260-4	8.196	7.100	28703017	25112945	276.361	264.050
35) L7 AR-1260-5	8.531	7.342	50695326	55968439	264.556	256.864



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092623\
Data File : PP060252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2023 17:39
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 18:14:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 26 17:51:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 58026550
Conc: 25.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 38553104
Conc: 25.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.456 min
Delta R.T.: 0.000 min
Response: 47066267
Conc: 27.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.751 min
Delta R.T.: 0.000 min
Response: 42803182
Conc: 26.32 ng/ml



#3 AR-1016-1


R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 19807608
Conc: 272.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 15556289
Conc: 269.20 ng/ml




#4 AR-1016-2

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 27770671
Conc: 268.15 ng/ml



#4 AR-1016-2


R.T.: 4.803 min
Delta R.T.: 0.000 min
Response: 21292150
Conc: 264.28 ng/ml



#5 AR-1016-3


R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 17819580
Conc: 274.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 4.981 min
Delta R.T.: 0.000 min
Response: 12044638
Conc: 268.19 ng/ml




#6 AR-1016-4

R.T.: 5.906 min
Delta R.T.: 0.000 min
Response: 14771441
Conc: 272.57 ng/ml



#6 AR-1016-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 9787366
Conc: 272.86 ng/ml



#7 AR-1016-5


R.T.: 6.204 min
Delta R.T.: 0.000 min
Response: 15303079
Conc: 275.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 13058864
Conc: 271.01 ng/ml




#31 AR-1260-1

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 28323067
Conc: 276.47 ng/ml



#31 AR-1260-1


R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 26771388
Conc: 268.61 ng/ml



#32 AR-1260-2


R.T.: 7.602 min
Delta R.T.: 0.000 min
Response: 32370783
Conc: 276.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 31426517
Conc: 266.16 ng/ml




#33 AR-1260-3

R.T.: 7.965 min
Delta R.T.: 0.000 min
Response: 25220660
Conc: 278.28 ng/ml



#33 AR-1260-3


R.T.: 6.625 min
Delta R.T.: 0.000 min
Response: 30065921
Conc: 263.39 ng/ml



#34 AR-1260-4


R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 28703017
Conc: 276.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.100 min
Delta R.T.: 0.000 min
Response: 25112945
Conc: 264.05 ng/ml



#35 AR-1260-5

R.T.: 8.531 min
Delta R.T.: 0.000 min
Response: 50695326
Conc: 264.56 ng/ml



#35 AR-1260-5

R.T.: 7.342 min
Delta R.T.: 0.000 min
Response: 55968439
Conc: 256.86 ng/ml