

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092623\
 Data File : PP060255.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2023 18:28
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 03:10:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 03:08:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.549	3.685	108.4E6	72734878	50.000	50.000
2) SA Decachlor...	10.472	8.750	83397298	79311754	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.926	4.067	25169042	18194746	500.000	500.000
12) L3 AR-1232-2	5.464	4.804	13691392	18234177	500.000	500.000
13) L3 AR-1232-3	5.759	4.982	23208600	10007629	500.000	500.000
14) L3 AR-1232-4	5.922	5.067	12289804	9332118	500.000	500.000
15) L3 AR-1232-5	6.014	5.240	10360833	10399683	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092623\
Data File : PP060255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2023 18:28
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 03:10:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 03:08:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.549 min
Delta R.T.: 0.000 min
Response: 108374162
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 72734878
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.472 min
Delta R.T.: 0.000 min
Response: 83397298
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.750 min
Delta R.T.: 0.000 min
Response: 79311754
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 25169042
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.067 min
Delta R.T.: 0.000 min
Response: 18194746
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.464 min
Delta R.T.: 0.000 min
Response: 13691392
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 18234177
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 23208600
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 10007629
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.922 min
Delta R.T.: 0.000 min
Response: 12289804
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.067 min
Delta R.T.: 0.000 min
Response: 9332118
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 6.014 min
Delta R.T.: 0.000 min
Response: 10360833
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.240 min
Delta R.T.: 0.000 min
Response: 10399683
Conc: 500.00 ng/ml