

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092722\
 Data File : PP051694.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2022 16:14
 Operator : YP\AJ
 Sample : AR1268CCC500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/29/2022
 Supervised By :Ankita Jodhani 09/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 16:46:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.683	86581118	74167652	48.722	49.100
2) SA Decachlor...	10.256	8.782	108.3E6	99052912	70.058	82.667
Target Compounds						
41) L9 AR-1268-1	8.705	7.645	89090498	85625206	442.288	454.231m
42) L9 AR-1268-2	8.802	7.710	81404518	78235497	443.683	467.115
43) L9 AR-1268-3	9.037	7.920	69746682	66932363	432.252	467.286
44) L9 AR-1268-4	9.474	8.212	30544021	27333170	435.522	482.362
45) L9 AR-1268-5	9.903	8.514	216.4E6	202.8E6	439.351	474.193

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092722\
Data File : PP051694.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 16:14
Operator : YP\AJ
Sample : AR1268CCC500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/29/2022
Supervised By :Ankita Jodhani 09/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 16:46:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

