

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092722\
 Data File : PP051677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2022 09:54
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 11:41:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.679	91573952	78684325	51.531	52.090
2) SA Decachlor...	10.254	8.780	69293192	60039271	44.837	50.107
Target Compounds						
21) L5 AR-1248-1	5.602	4.783	19667424	16337034	495.934	505.306
22) L5 AR-1248-2	5.877	5.024	28906486	22115302	509.595	487.583
23) L5 AR-1248-3	6.083	5.066	31761215	23501220	505.858	497.173
24) L5 AR-1248-4	6.489	5.241	27875385	26565122	459.707	476.041
25) L5 AR-1248-5	6.528	5.641	27603511	26008953	442.275	493.365



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092722\
Data File : PP051677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 09:54
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 11:41:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: -0.004 min
Response: 91573952
Conc: 51.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.679 min
Delta R.T.: -0.005 min
Response: 78684325
Conc: 52.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.254 min
Delta R.T.: 0.000 min
Response: 69293192
Conc: 44.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.780 min
Delta R.T.: -0.008 min
Response: 60039271
Conc: 50.11 ng/ml



#21 AR-1248-1


R.T.: 5.602 min
Delta R.T.: -0.005 min
Response: 19667424
Conc: 495.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.783 min
Delta R.T.: -0.007 min
Response: 16337034
Conc: 505.31 ng/ml




#22 AR-1248-2

R.T.: 5.877 min
Delta R.T.: -0.004 min
Response: 28906486
Conc: 509.59 ng/ml



#22 AR-1248-2


R.T.: 5.024 min
Delta R.T.: -0.007 min
Response: 22115302
Conc: 487.58 ng/ml



#23 AR-1248-3


R.T.: 6.083 min
Delta R.T.: -0.005 min
Response: 31761215
Conc: 505.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.066 min
Delta R.T.: -0.008 min
Response: 23501220
Conc: 497.17 ng/ml




#24 AR-1248-4

R.T.: 6.489 min
Delta R.T.: -0.004 min
Response: 27875385
Conc: 459.71 ng/ml



#24 AR-1248-4


R.T.: 5.241 min
Delta R.T.: -0.008 min
Response: 26565122
Conc: 476.04 ng/ml



#25 AR-1248-5

R.T.: 6.528 min
Delta R.T.: -0.004 min
Response: 27603511
Conc: 442.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.641 min
Delta R.T.: -0.008 min
Response: 26008953
Conc: 493.37 ng/ml