

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
 Data File : PP060299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 13:55
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 14:42:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.532	3.684	109.3E6	73555126	48.797	49.415
2)	SA Decachlor...	10.457	8.750	83105595	77287486	47.071	47.381
Target Compounds							
3)	L1 AR-1016-1	5.718	4.783	34646954	28310940	470.717	487.198
4)	L1 AR-1016-2	5.741	4.802	49500329	38730258	472.633	481.088
5)	L1 AR-1016-3	5.805	4.980	30721036	21639734	470.054	484.191
6)	L1 AR-1016-4	5.905	5.022	25555462	16921810	469.947	470.263
7)	L1 AR-1016-5	6.203	5.237	25553650	23114353	459.410	475.702
8)	L2 AR-1221-1	4.738	3.900	4725321	3358122	163.580	160.752
9)	L2 AR-1221-2	4.830	3.988	6455502	4193304	302.153	281.224
10)	L2 AR-1221-3	4.907	4.065	20663687	14788317	340.261	336.651
11)	L3 AR-1232-1	4.907f	4.065	20663687	14788317	410.498	406.390
12)	L3 AR-1232-2	5.447f	4.802	27200898	38730258	993.358	1062.024
13)	L3 AR-1232-3	5.741f	4.980	49500329	21639734	1066.422	1081.162
14)	L3 AR-1232-4	5.905f	5.064	25555462	21216398	1039.702	1136.741
15)	L3 AR-1232-5	5.996f	5.237	22595841	23114353	1090.445	1111.301
16)	L4 AR-1242-1	5.718	4.783	34646954	28310940	542.145	559.205
17)	L4 AR-1242-2	5.741	4.802	49500329	38730258	546.685	557.716
18)	L4 AR-1242-3	5.805	4.980	30721036	21639734	545.661	561.915
19)	L4 AR-1242-4	5.905	5.064	25555462	21216398	547.157	538.721
20)	L4 AR-1242-5	6.651	5.593	7603382	22665100	151.173	463.336 #
21)	L5 AR-1248-1	5.718	4.783	34646954	28310940	744.704	775.778
22)	L5 AR-1248-2	5.996	5.022	22595841	16921810	325.893	323.727
23)	L5 AR-1248-3	6.203	5.064	25553650	21216398	336.455	382.110
24)	L5 AR-1248-4	6.610	5.237	8344096	23114353	100.889	351.137 #
25)	L5 AR-1248-5	6.651	5.634	7603382	5830699	94.472	90.648
26)	L6 AR-1254-1	6.585	5.593	22903952	22665100	268.384	229.128
27)	L6 AR-1254-2	6.806	5.741	19720316	16414796	154.516	190.996
28)	L6 AR-1254-3	7.176	6.164	9400513	31588444	70.725	220.395 #
29)	L6 AR-1254-4	7.465	6.378	6643386	2490341	67.556	28.571 #
30)	L6 AR-1254-5	7.886	6.798	63493897	60073343	586.044	474.063
31)	L7 AR-1260-1	7.342	6.280	47329783	46207885	451.513	460.351
32)	L7 AR-1260-2	7.601	6.468	55177550	55237805	456.367	462.998
33)	L7 AR-1260-3	7.964	6.623	42538498	52809974	458.932	461.245
34)	L7 AR-1260-4	8.196	7.099	48693032	43745193	459.148	459.596
35)	L7 AR-1260-5	8.530	7.341	89953948	100.3E6	464.828	463.013
36)	L8 AR-1262-1	7.964	6.891	42538498	23255913	341.030	434.305 #
37)	L8 AR-1262-2	8.530	7.099	89953948	43745193	429.409	370.879
38)	L8 AR-1262-3	8.872	7.627	60435843	22151306	399.345	240.230 #
39)	L8 AR-1262-4	8.935f	7.690	37966117	70821717	317.240	420.402 #
40)	L8 AR-1262-5	9.649	8.190	25814766	24820400	322.426	320.369
41)	L9 AR-1268-1	8.872f	7.627	60435843	22151306	227.802	81.492 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
 Data File : PP060299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 13:55
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 14:42:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.935f	7.690	37966117	70821717	155.768	290.403 #
43) L9	AR-1268-3	9.198	7.901	1568040	1446145	7.497	6.761
44) L9	AR-1268-4	9.649	8.190	25814766	24820400	279.433	282.892
45) L9	AR-1268-5	10.093	8.488	7507618	6951620	11.858	11.122



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
Data File : PP060299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 13:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 14:42:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 109281772
Conc: 48.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 73555126
Conc: 49.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.457 min
Delta R.T.: 0.001 min
Response: 83105595
Conc: 47.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.750 min
Delta R.T.: 0.000 min
Response: 77287486
Conc: 47.38 ng/ml



#3 AR-1016-1


R.T.: 5.718 min
Delta R.T.: -0.001 min
Response: 34646954
Conc: 470.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 28310940
Conc: 487.20 ng/ml




#4 AR-1016-2

R.T.: 5.741 min
Delta R.T.: 0.000 min
Response: 49500329
Conc: 472.63 ng/ml



#4 AR-1016-2


R.T.: 4.802 min
Delta R.T.: -0.002 min
Response: 38730258
Conc: 481.09 ng/ml



#5 AR-1016-3


R.T.: 5.805 min
Delta R.T.: -0.001 min
Response: 30721036
Conc: 470.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 21639734
Conc: 484.19 ng/ml




#6 AR-1016-4

R.T.: 5.905 min
Delta R.T.: -0.001 min
Response: 25555462
Conc: 469.95 ng/ml



#6 AR-1016-4


R.T.: 5.022 min
Delta R.T.: -0.002 min
Response: 16921810
Conc: 470.26 ng/ml



#7 AR-1016-5


R.T.: 6.203 min
Delta R.T.: 0.000 min
Response: 25553650
Conc: 459.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 23114353
Conc: 475.70 ng/ml




#8 AR-1221-1

R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 4725321
Conc: 163.58 ng/ml



#8 AR-1221-1


R.T.: 3.900 min
Delta R.T.: -0.004 min
Response: 3358122
Conc: 160.75 ng/ml



#9 AR-1221-2


R.T.: 4.830 min
Delta R.T.: 0.000 min
Response: 6455502
Conc: 302.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.988 min
Delta R.T.: -0.001 min
Response: 4193304
Conc: 281.22 ng/ml




#10 AR-1221-3

R.T.: 4.907 min
Delta R.T.: -0.002 min
Response: 20663687
Conc: 340.26 ng/ml



#10 AR-1221-3


R.T.: 4.065 min
Delta R.T.: -0.002 min
Response: 14788317
Conc: 336.65 ng/ml



#11 AR-1232-1


R.T.: 4.907 min
Delta R.T.: -0.019 min
Response: 20663687
Conc: 410.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.065 min
Delta R.T.: -0.002 min
Response: 14788317
Conc: 406.39 ng/ml




#12 AR-1232-2

R.T.: 5.447 min
Delta R.T.: -0.017 min
Response: 27200898
Conc: 993.36 ng/ml



#12 AR-1232-2


R.T.: 4.802 min
Delta R.T.: -0.002 min
Response: 38730258
Conc: 1062.02 ng/ml



#13 AR-1232-3


R.T.: 5.741 min
Delta R.T.: -0.018 min
Response: 49500329
Conc: 1066.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 21639734
Conc: 1081.16 ng/ml




#14 AR-1232-4

R.T.: 5.905 min
Delta R.T.: -0.018 min
Response: 25555462
Conc: 1039.70 ng/ml



#14 AR-1232-4


R.T.: 5.064 min
Delta R.T.: -0.002 min
Response: 21216398
Conc: 1136.74 ng/ml



#15 AR-1232-5


R.T.: 5.996 min
Delta R.T.: -0.017 min
Response: 22595841
Conc: 1090.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 23114353
Conc: 1111.30 ng/ml




#16 AR-1242-1

R.T.: 5.718 min
Delta R.T.: 0.000 min
Response: 34646954
Conc: 542.15 ng/ml



#16 AR-1242-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 28310940
Conc: 559.20 ng/ml



#17 AR-1242-2


R.T.: 5.741 min
Delta R.T.: 0.000 min
Response: 49500329
Conc: 546.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.802 min
Delta R.T.: -0.001 min
Response: 38730258
Conc: 557.72 ng/ml




#18 AR-1242-3

R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 30721036
Conc: 545.66 ng/ml



#18 AR-1242-3


R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 21639734
Conc: 561.91 ng/ml



#19 AR-1242-4


R.T.: 5.905 min
Delta R.T.: 0.000 min
Response: 25555462
Conc: 547.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.064 min
Delta R.T.: -0.002 min
Response: 21216398
Conc: 538.72 ng/ml




#20 AR-1242-5

R.T.: 6.651 min
Delta R.T.: 0.000 min
Response: 7603382
Conc: 151.17 ng/ml



#20 AR-1242-5


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 22665100
Conc: 463.34 ng/ml



#21 AR-1248-1


R.T.: 5.718 min
Delta R.T.: 0.000 min
Response: 34646954
Conc: 744.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.783 min
Delta R.T.: 0.000 min
Response: 28310940
Conc: 775.78 ng/ml




#22 AR-1248-2

R.T.: 5.996 min
Delta R.T.: 0.000 min
Response: 22595841
Conc: 325.89 ng/ml



#22 AR-1248-2


R.T.: 5.022 min
Delta R.T.: 0.000 min
Response: 16921810
Conc: 323.73 ng/ml



#23 AR-1248-3


R.T.: 6.203 min
Delta R.T.: 0.000 min
Response: 25553650
Conc: 336.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.064 min
Delta R.T.: 0.000 min
Response: 21216398
Conc: 382.11 ng/ml




#24 AR-1248-4

R.T.: 6.610 min
Delta R.T.: -0.002 min
Response: 8344096
Conc: 100.89 ng/ml



#24 AR-1248-4


R.T.: 5.237 min
Delta R.T.: -0.001 min
Response: 23114353
Conc: 351.14 ng/ml



#25 AR-1248-5


R.T.: 6.651 min
Delta R.T.: 0.000 min
Response: 7603382
Conc: 94.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.634 min
Delta R.T.: -0.001 min
Response: 5830699
Conc: 90.65 ng/ml




#26 AR-1254-1

R.T.: 6.585 min
Delta R.T.: 0.000 min
Response: 22903952
Conc: 268.38 ng/ml



#26 AR-1254-1


R.T.: 5.593 min
Delta R.T.: -0.001 min
Response: 22665100
Conc: 229.13 ng/ml



#27 AR-1254-2


R.T.: 6.806 min
Delta R.T.: 0.000 min
Response: 19720316
Conc: 154.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.741 min
Delta R.T.: 0.000 min
Response: 16414796
Conc: 191.00 ng/ml




#28 AR-1254-3

R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 9400513
Conc: 70.73 ng/ml



#28 AR-1254-3


R.T.: 6.164 min
Delta R.T.: 0.015 min
Response: 31588444
Conc: 220.39 ng/ml



#29 AR-1254-4


R.T.: 7.465 min
Delta R.T.: 0.000 min
Response: 6643386
Conc: 67.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 2490341
Conc: 28.57 ng/ml




#30 AR-1254-5

R.T.: 7.886 min
Delta R.T.: 0.000 min
Response: 63493897
Conc: 586.04 ng/ml



#30 AR-1254-5


R.T.: 6.798 min
Delta R.T.: 0.000 min
Response: 60073343
Conc: 474.06 ng/ml



#31 AR-1260-1


R.T.: 7.342 min
Delta R.T.: -0.001 min
Response: 47329783
Conc: 451.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: -0.002 min
Response: 46207885
Conc: 460.35 ng/ml




#32 AR-1260-2

R.T.: 7.601 min
Delta R.T.: 0.000 min
Response: 55177550
Conc: 456.37 ng/ml



#32 AR-1260-2


R.T.: 6.468 min
Delta R.T.: -0.002 min
Response: 55237805
Conc: 463.00 ng/ml



#33 AR-1260-3


R.T.: 7.964 min
Delta R.T.: -0.001 min
Response: 42538498
Conc: 458.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.623 min
Delta R.T.: -0.002 min
Response: 52809974
Conc: 461.24 ng/ml




#34 AR-1260-4

R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 48693032
Conc: 459.15 ng/ml



#34 AR-1260-4


R.T.: 7.099 min
Delta R.T.: -0.002 min
Response: 43745193
Conc: 459.60 ng/ml



#35 AR-1260-5


R.T.: 8.530 min
Delta R.T.: 0.000 min
Response: 89953948
Conc: 464.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.341 min
Delta R.T.: 0.000 min
Response: 100341388
Conc: 463.01 ng/ml




#36 AR-1262-1

R.T.: 7.964 min
Delta R.T.: 0.000 min
Response: 42538498
Conc: 341.03 ng/ml



#36 AR-1262-1


R.T.: 6.891 min
Delta R.T.: 0.000 min
Response: 23255913
Conc: 434.30 ng/ml



#37 AR-1262-2


R.T.: 8.530 min
Delta R.T.: 0.001 min
Response: 89953948
Conc: 429.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 43745193
Conc: 370.88 ng/ml




#38 AR-1262-3

R.T.: 8.872 min
Delta R.T.: 0.012 min
Response: 60435843
Conc: 399.35 ng/ml



#38 AR-1262-3


R.T.: 7.627 min
Delta R.T.: 0.000 min
Response: 22151306
Conc: 240.23 ng/ml



#39 AR-1262-4


R.T.: 8.935 min
Delta R.T.: -0.017 min
Response: 37966117
Conc: 317.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 70821717
Conc: 420.40 ng/ml




#40 AR-1262-5

R.T.: 9.649 min
Delta R.T.: 0.003 min
Response: 25814766
Conc: 322.43 ng/ml



#40 AR-1262-5


R.T.: 8.190 min
Delta R.T.: 0.000 min
Response: 24820400
Conc: 320.37 ng/ml



#41 AR-1268-1


R.T.: 8.872 min
Delta R.T.: 0.015 min
Response: 60435843
Conc: 227.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.627 min
Delta R.T.: 0.000 min
Response: 22151306
Conc: 81.49 ng/ml




#42 AR-1268-2

R.T.: 8.935 min
Delta R.T.: -0.021 min
Response: 37966117
Conc: 155.77 ng/ml



#42 AR-1268-2


R.T.: 7.690 min
Delta R.T.: -0.001 min
Response: 70821717
Conc: 290.40 ng/ml



#43 AR-1268-3


R.T.: 9.198 min
Delta R.T.: 0.000 min
Response: 1568040
Conc: 7.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.901 min
Delta R.T.: 0.000 min
Response: 1446145
Conc: 6.76 ng/ml




#44 AR-1268-4

R.T.: 9.649 min
Delta R.T.: 0.001 min
Response: 25814766
Conc: 279.43 ng/ml



#44 AR-1268-4


R.T.: 8.190 min
Delta R.T.: 0.000 min
Response: 24820400
Conc: 282.89 ng/ml



#45 AR-1268-5

R.T.: 10.093 min
Delta R.T.: 0.002 min
Response: 7507618
Conc: 11.86 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.488 min
Delta R.T.: 0.001 min
Response: 6951620
Conc: 11.12 ng/ml