

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
 Data File : PP060308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 16:36
 Operator : YP\AJ
 Sample : 04615-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AUD-40281

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/28/2023
 Supervised By :Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:19:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.533	3.684	48008316	35908199	21.437	24.124m
2) SA Decachlor...	10.459	8.751	40955499	36376085	23.197	22.300

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
Data File : PP060308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 16:36
Operator : YP\AJ
Sample : 04615-03
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AUD-40281


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/28/2023
Supervised By :Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:19:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 48008316
Conc: 21.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AUD-40281


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/28/2023
Supervised By :Ankita Jodhani 09/28/2023




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 35908199
Conc: 24.12 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.459 min
Delta R.T.: 0.004 min
Response: 40955499
Conc: 23.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.751 min
Delta R.T.: 0.000 min
Response: 36376085
Conc: 22.30 ng/ml