

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
 Data File : PP060325.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2023 21:54
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 28 01:24:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	114.2E6	83843892	50.998	56.328m
2) SA Decachlor...	10.456	8.749	89074034	81950029	50.452	50.239
Target Compounds						
26) L6 AR-1254-1	6.586	5.593	42649898	49050603	499.763	495.867
27) L6 AR-1254-2	6.807	5.742	63804251	42083142	499.930	489.664
28) L6 AR-1254-3	7.177	6.149	64993369	69813597	488.980	487.094
29) L6 AR-1254-4	7.465	6.378	47658590	42145649	484.635	483.527
30) L6 AR-1254-5	7.888	6.799	52264105	61262542	482.394	483.447



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092723\
Data File : PP060325.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 21:54
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 28 01:24:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 114209198
Conc: 51.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 83843892
Conc: 56.33 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.456 min
Delta R.T.: 0.000 min
Response: 89074034
Conc: 50.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.749 min
Delta R.T.: -0.001 min
Response: 81950029
Conc: 50.24 ng/ml



#26 AR-1254-1


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 42649898
Conc: 499.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 49050603
Conc: 495.87 ng/ml




#27 AR-1254-2

R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 63804251
Conc: 499.93 ng/ml



#27 AR-1254-2


R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 42083142
Conc: 489.66 ng/ml



#28 AR-1254-3


R.T.: 7.177 min
Delta R.T.: 0.000 min
Response: 64993369
Conc: 488.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.149 min
Delta R.T.: 0.000 min
Response: 69813597
Conc: 487.09 ng/ml




#29 AR-1254-4

R.T.: 7.465 min
Delta R.T.: 0.000 min
Response: 47658590
Conc: 484.64 ng/ml



#29 AR-1254-4


R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 42145649
Conc: 483.53 ng/ml



#30 AR-1254-5

R.T.: 7.888 min
Delta R.T.: 0.000 min
Response: 52264105
Conc: 482.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.799 min
Delta R.T.: 0.000 min
Response: 61262542
Conc: 483.45 ng/ml