

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092821\
 Data File : PP039594.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Sep 2021 21:24
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 05:06:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.907	3.997	1732054	925964	54.712	56.828
2) SA Decachlor...	10.931	9.336	1157292	1166498	53.346	53.381
Target Compounds						
26) L6 AR-1254-1	7.140	6.140	595537	566870	545.608	544.476
27) L6 AR-1254-2	7.370	6.299	850509	513578	526.546	532.920
28) L6 AR-1254-3	7.752	6.723	853117	851781	512.733	566.212
29) L6 AR-1254-4	8.049	6.963	613821	524013	502.500	544.182
30) L6 AR-1254-5	8.479	7.395	623083	765402	491.080	543.789



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092821\
Data File : PP039594.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Sep 2021 21:24
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 05:06:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 25 01:28:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.907 min
Delta R.T.: 0.000 min
Response: 1732054
Conc: 54.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.997 min
Delta R.T.: 0.000 min
Response: 925964
Conc: 56.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.931 min
Delta R.T.: 0.002 min
Response: 1157292
Conc: 53.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.336 min
Delta R.T.: 0.000 min
Response: 1166498
Conc: 53.38 ng/ml



#26 AR-1254-1


R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 595537
Conc: 545.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 566870
Conc: 544.48 ng/ml




#27 AR-1254-2

R.T.: 7.370 min
Delta R.T.: 0.000 min
Response: 850509
Conc: 526.55 ng/ml



#27 AR-1254-2


R.T.: 6.299 min
Delta R.T.: -0.002 min
Response: 513578
Conc: 532.92 ng/ml



#28 AR-1254-3


R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 853117
Conc: 512.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.723 min
Delta R.T.: -0.001 min
Response: 851781
Conc: 566.21 ng/ml




#29 AR-1254-4

R.T.: 8.049 min
Delta R.T.: 0.000 min
Response: 613821
Conc: 502.50 ng/ml



#29 AR-1254-4


R.T.: 6.963 min
Delta R.T.: -0.001 min
Response: 524013
Conc: 544.18 ng/ml



#30 AR-1254-5

R.T.: 8.479 min
Delta R.T.: 0.000 min
Response: 623083
Conc: 491.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.395 min
Delta R.T.: 0.000 min
Response: 765402
Conc: 543.79 ng/ml