

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092822\
 Data File : PP051762.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2022 17:00
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 00:35:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.431	3.682	86490864	75726575	48.671	50.132
2) SA Decachlor...	10.252	8.779	70743580	54469569	45.775	45.459
Target Compounds						
21) L5 AR-1248-1	5.604	4.785	18638528	15445131	469.989	477.719
22) L5 AR-1248-2	5.879	5.025	26623172	21224013	469.342	467.932
23) L5 AR-1248-3	6.085	5.068	29437542	22214792	468.849	469.958
24) L5 AR-1248-4	6.490	5.243	26154808	25703361	431.332	460.598
25) L5 AR-1248-5	6.529	5.642	25902984	24134619	415.028	457.811



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092822\
Data File : PP051762.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2022 17:00
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 00:35:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.431 min
Delta R.T.: -0.001 min
Response: 86490864
Conc: 48.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.002 min
Response: 75726575
Conc: 50.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.252 min
Delta R.T.: -0.001 min
Response: 70743580
Conc: 45.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.779 min
Delta R.T.: -0.008 min
Response: 54469569
Conc: 45.46 ng/ml



#21 AR-1248-1


R.T.: 5.604 min
Delta R.T.: -0.002 min
Response: 18638528
Conc: 469.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.785 min
Delta R.T.: -0.005 min
Response: 15445131
Conc: 477.72 ng/ml




#22 AR-1248-2

R.T.: 5.879 min
Delta R.T.: -0.003 min
Response: 26623172
Conc: 469.34 ng/ml



#22 AR-1248-2


R.T.: 5.025 min
Delta R.T.: -0.006 min
Response: 21224013
Conc: 467.93 ng/ml



#23 AR-1248-3


R.T.: 6.085 min
Delta R.T.: -0.003 min
Response: 29437542
Conc: 468.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.068 min
Delta R.T.: -0.006 min
Response: 22214792
Conc: 469.96 ng/ml




#24 AR-1248-4

R.T.: 6.490 min
Delta R.T.: -0.003 min
Response: 26154808
Conc: 431.33 ng/ml



#24 AR-1248-4


R.T.: 5.243 min
Delta R.T.: -0.006 min
Response: 25703361
Conc: 460.60 ng/ml



#25 AR-1248-5

R.T.: 6.529 min
Delta R.T.: -0.003 min
Response: 25902984
Conc: 415.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.642 min
Delta R.T.: -0.006 min
Response: 24134619
Conc: 457.81 ng/ml