

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092824\
 Data File : PP067303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2024 14:45
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/30/2024
 Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 01:30:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.753	4.052	62621950	68537084	46.590	45.870
2) SA Decachlor...	10.664	9.225	59782463	58881459	41.181	42.729
Target Compounds						
26) L6 AR-1254-1	6.772	5.979	26574207	35080249	426.529m	432.763
27) L6 AR-1254-2	6.990	6.126	38554127	31578150	424.912	430.557
28) L6 AR-1254-3	7.355	6.536	39652975	48502853	422.746	425.393
29) L6 AR-1254-4	7.637	6.764	27751003	26600649	416.407	426.189
30) L6 AR-1254-5	8.055	7.186	33812275	43765690	411.362m	426.277

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092824\
Data File : PP067303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2024 14:45
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 01:30:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.753 min
Delta R.T.: 0.000 min
Response: 62621950
Conc: 46.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: -0.001 min
Response: 68537084
Conc: 45.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.664 min
Delta R.T.: 0.002 min
Response: 59782463
Conc: 41.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.225 min
Delta R.T.: -0.001 min
Response: 58881459
Conc: 42.73 ng/ml




#26 AR-1254-1

R.T.: 6.772 min
Delta R.T.: 0.002 min
Response: 26574207
Conc: 426.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024




#26 AR-1254-1

R.T.: 5.979 min
Delta R.T.: -0.001 min
Response: 35080249
Conc: 432.76 ng/ml




#27 AR-1254-2

R.T.: 6.990 min
Delta R.T.: 0.002 min
Response: 38554127
Conc: 424.91 ng/ml



#27 AR-1254-2

R.T.: 6.126 min
Delta R.T.: -0.001 min
Response: 31578150
Conc: 430.56 ng/ml




#28 AR-1254-3

R.T.: 7.355 min
Delta R.T.: 0.002 min
Response: 39652975
Conc: 422.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024




#28 AR-1254-3

R.T.: 6.536 min
Delta R.T.: -0.001 min
Response: 48502853
Conc: 425.39 ng/ml




#29 AR-1254-4

R.T.: 7.637 min
Delta R.T.: 0.002 min
Response: 27751003
Conc: 416.41 ng/ml



#29 AR-1254-4

R.T.: 6.764 min
Delta R.T.: -0.001 min
Response: 26600649
Conc: 426.19 ng/ml




#30 AR-1254-5

R.T.: 8.055 min
Delta R.T.: 0.003 min
Response: 33812275
Conc: 411.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024



#30 AR-1254-5

R.T.: 7.186 min
Delta R.T.: -0.002 min
Response: 43765690
Conc: 426.28 ng/ml