

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092824\
 Data File : PP067307.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2024 15:49
 Operator : YP\AJ
 Sample : P4217-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 RV-22(0-0.5)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/30/2024
 Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 01:31:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.754	4.053	22620265	24650715	16.829	16.498
2) SA Decachlor...	10.664	9.224	28986175	24384220	19.967m	17.695m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092824\
Data File : PP067307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2024 15:49
Operator : YP\AJ
Sample : P4217-02
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
RV-22(0-0.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 01:31:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.754 min
Delta R.T.: 0.000 min
Response: 22620265
Conc: 16.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
RV-22(0-0.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2024
Supervised By :Ankita Jodhani 09/30/2024




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: 0.000 min
Response: 24650715
Conc: 16.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.664 min
Delta R.T.: 0.003 min
Response: 28986175
Conc: 19.97 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.224 min
Delta R.T.: -0.002 min
Response: 24384220
Conc: 17.70 ng/ml m