

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092921\
 Data File : PP039641.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Sep 2021 1:16
 Operator : AJ\MA
 Sample : M3976-05MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 04:01:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.906	3.995	734843	403425	23.212	24.759
2)	SA Decachlor...	10.929	9.331	479816	504014	22.118	23.065
Target Compounds							
3)	L1 AR-1016-1	6.228	5.259	612669	415760	571.424	576.896
4)	L1 AR-1016-2	6.252	5.280	893220	579916	569.412	587.749
5)	L1 AR-1016-3	6.317	5.473	556813	323964	568.651	594.228
6)	L1 AR-1016-4	6.425	5.525	441791	254844	564.237	575.154
7)	L1 AR-1016-5	6.740	5.756	428019	324015	578.287	586.091
8)	L2 AR-1221-1	5.154	4.255	63123	43675	168.191	216.778 #
9)	L2 AR-1221-2	5.255	4.356	75062	51280	275.724	300.917
10)	L2 AR-1221-3	5.341	4.445	355412	218192	415.182	403.553
11)	L3 AR-1232-1	5.341	4.445	355412	218192	555.334	539.554
12)	L3 AR-1232-2	5.932	5.280	468215	579916	1300.630	1388.709
13)	L3 AR-1232-3	6.252	5.473	893220	323964	1372.747	1460.275
14)	L3 AR-1232-4	6.425	5.570	441791	319468	1318.002	1361.170
15)	L3 AR-1232-5	6.522	5.756	355784	324015	1543.726	1483.241
16)	L4 AR-1242-1	6.228	5.259	612669	415760	757.029	775.133
17)	L4 AR-1242-2	6.252	5.280	893220	579916	765.753	791.807
18)	L4 AR-1242-3	6.317	5.473	556813	323964	759.704	786.759
19)	L4 AR-1242-4	6.425	5.570	441791	319468	738.687	756.243
20)	L4 AR-1242-5	7.210	6.138	90460	322257	145.650	607.285 #
21)	L5 AR-1248-1	6.228	5.259	612669	415760	954.091	957.369
22)	L5 AR-1248-2	6.522	5.525	355784	254844	431.166	430.047
23)	L5 AR-1248-3	6.740	5.570	428019	319468	435.168	527.244
24)	L5 AR-1248-4	7.169	5.756	96960	324015	89.317	475.839 #
25)	L5 AR-1248-5	7.210	6.180	90460	50009	86.689	70.948
26)	L6 AR-1254-1	7.140	6.138	409090	322257	374.792	309.526
27)	L6 AR-1254-2	7.369	6.297	426888	258588	264.284	268.328
28)	L6 AR-1254-3	7.752	6.738	234924	457729	141.192	304.270 #
29)	L6 AR-1254-4	8.045	6.960	123793	56721	101.342	58.904 #
30)	L6 AR-1254-5	8.476	7.391	987983	889816	778.674	632.181
31)	L7 AR-1260-1	7.918	6.858	766288	629813	623.570	584.833
32)	L7 AR-1260-2	8.184	7.055	877380	759300	633.747	592.835
33)	L7 AR-1260-3	8.549	7.212	493784	708260	596.996	588.725
34)	L7 AR-1260-4	8.781	7.697	597929	537151	599.749	576.755
35)	L7 AR-1260-5	9.106	7.945	1073602	1201743	566.549	560.271
36)	L8 AR-1262-1	8.549	7.391	493784	889816	369.799	1262.175 #
37)	L8 AR-1262-2	9.106	7.945	1073602	1201743	488.099	528.997
38)	L8 AR-1262-3	9.439f	8.234	726084	253515	743.794	249.429 #
39)	L8 AR-1262-4	9.490f	8.297	410354	898205	686.038	500.060 #
40)	L8 AR-1262-5	10.176	8.797	266608	300658	342.444	330.228
41)	L9 AR-1268-1	9.439f	8.234	726084	253515	269.628	86.943 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092921\
 Data File : PP039641.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Sep 2021 1:16
 Operator : AJ\MA
 Sample : M3976-05MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 04:01:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.490f	8.297	410354	898205	159.309	336.949 #
43) L9	AR-1268-3	0.000	8.509	0	21287	N.D.	9.278 #
44) L9	AR-1268-4	10.176	8.797	266608	300658	290.217	292.102
45) L9	AR-1268-5	10.591	9.081	95294	99541	12.177	13.565



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092921\
Data File : PP039641.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Sep 2021 1:16
Operator : AJ\MA
Sample : M3976-05MSD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 04:01:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 25 01:28:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 734843
Conc: 23.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 403425
Conc: 24.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.929 min
Delta R.T.: -0.001 min
Response: 479816
Conc: 22.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.331 min
Delta R.T.: -0.005 min
Response: 504014
Conc: 23.06 ng/ml



#3 AR-1016-1


R.T.: 6.228 min
Delta R.T.: -0.001 min
Response: 612669
Conc: 571.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.259 min
Delta R.T.: -0.003 min
Response: 415760
Conc: 576.90 ng/ml




#4 AR-1016-2

R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 893220
Conc: 569.41 ng/ml



#4 AR-1016-2


R.T.: 5.280 min
Delta R.T.: -0.003 min
Response: 579916
Conc: 587.75 ng/ml



#5 AR-1016-3


R.T.: 6.317 min
Delta R.T.: -0.001 min
Response: 556813
Conc: 568.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.473 min
Delta R.T.: -0.003 min
Response: 323964
Conc: 594.23 ng/ml




#6 AR-1016-4

R.T.: 6.425 min
Delta R.T.: 0.000 min
Response: 441791
Conc: 564.24 ng/ml



#6 AR-1016-4


R.T.: 5.525 min
Delta R.T.: -0.003 min
Response: 254844
Conc: 575.15 ng/ml



#7 AR-1016-5


R.T.: 6.740 min
Delta R.T.: 0.000 min
Response: 428019
Conc: 578.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.756 min
Delta R.T.: -0.003 min
Response: 324015
Conc: 586.09 ng/ml




#8 AR-1221-1

R.T.: 5.154 min
Delta R.T.: -0.002 min
Response: 63123
Conc: 168.19 ng/ml



#8 AR-1221-1


R.T.: 4.255 min
Delta R.T.: -0.004 min
Response: 43675
Conc: 216.78 ng/ml



#9 AR-1221-2


R.T.: 5.255 min
Delta R.T.: 0.000 min
Response: 75062
Conc: 275.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.356 min
Delta R.T.: -0.002 min
Response: 51280
Conc: 300.92 ng/ml




#10 AR-1221-3

R.T.: 5.341 min
Delta R.T.: -0.001 min
Response: 355412
Conc: 415.18 ng/ml



#10 AR-1221-3


R.T.: 4.445 min
Delta R.T.: -0.003 min
Response: 218192
Conc: 403.55 ng/ml



#11 AR-1232-1


R.T.: 5.341 min
Delta R.T.: 0.000 min
Response: 355412
Conc: 555.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.445 min
Delta R.T.: -0.002 min
Response: 218192
Conc: 539.55 ng/ml




#12 AR-1232-2

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 468215
Conc: 1300.63 ng/ml



#12 AR-1232-2


R.T.: 5.280 min
Delta R.T.: -0.003 min
Response: 579916
Conc: 1388.71 ng/ml



#13 AR-1232-3


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 893220
Conc: 1372.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.473 min
Delta R.T.: -0.002 min
Response: 323964
Conc: 1460.27 ng/ml




#14 AR-1232-4

R.T.: 6.425 min
Delta R.T.: 0.000 min
Response: 441791
Conc: 1318.00 ng/ml



#14 AR-1232-4


R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 319468
Conc: 1361.17 ng/ml



#15 AR-1232-5


R.T.: 6.522 min
Delta R.T.: 0.000 min
Response: 355784
Conc: 1543.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.756 min
Delta R.T.: -0.003 min
Response: 324015
Conc: 1483.24 ng/ml




#16 AR-1242-1

R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 612669
Conc: 757.03 ng/ml



#16 AR-1242-1


R.T.: 5.259 min
Delta R.T.: -0.003 min
Response: 415760
Conc: 775.13 ng/ml



#17 AR-1242-2


R.T.: 6.252 min
Delta R.T.: 0.000 min
Response: 893220
Conc: 765.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.280 min
Delta R.T.: -0.003 min
Response: 579916
Conc: 791.81 ng/ml




#18 AR-1242-3

R.T.: 6.317 min
Delta R.T.: -0.001 min
Response: 556813
Conc: 759.70 ng/ml



#18 AR-1242-3


R.T.: 5.473 min
Delta R.T.: -0.003 min
Response: 323964
Conc: 786.76 ng/ml



#19 AR-1242-4


R.T.: 6.425 min
Delta R.T.: -0.001 min
Response: 441791
Conc: 738.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 319468
Conc: 756.24 ng/ml




#20 AR-1242-5

R.T.: 7.210 min
Delta R.T.: -0.001 min
Response: 90460
Conc: 145.65 ng/ml



#20 AR-1242-5


R.T.: 6.138 min
Delta R.T.: -0.002 min
Response: 322257
Conc: 607.28 ng/ml



#21 AR-1248-1


R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 612669
Conc: 954.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.259 min
Delta R.T.: -0.003 min
Response: 415760
Conc: 957.37 ng/ml




#22 AR-1248-2

R.T.: 6.522 min
Delta R.T.: -0.001 min
Response: 355784
Conc: 431.17 ng/ml



#22 AR-1248-2


R.T.: 5.525 min
Delta R.T.: -0.003 min
Response: 254844
Conc: 430.05 ng/ml



#23 AR-1248-3


R.T.: 6.740 min
Delta R.T.: -0.002 min
Response: 428019
Conc: 435.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 319468
Conc: 527.24 ng/ml




#24 AR-1248-4

R.T.: 7.169 min
Delta R.T.: -0.002 min
Response: 96960
Conc: 89.32 ng/ml



#24 AR-1248-4


R.T.: 5.756 min
Delta R.T.: -0.003 min
Response: 324015
Conc: 475.84 ng/ml



#25 AR-1248-5


R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 90460
Conc: 86.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.180 min
Delta R.T.: -0.003 min
Response: 50009
Conc: 70.95 ng/ml




#26 AR-1254-1

R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 409090
Conc: 374.79 ng/ml



#26 AR-1254-1


R.T.: 6.138 min
Delta R.T.: -0.002 min
Response: 322257
Conc: 309.53 ng/ml



#27 AR-1254-2


R.T.: 7.369 min
Delta R.T.: -0.002 min
Response: 426888
Conc: 264.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.297 min
Delta R.T.: -0.004 min
Response: 258588
Conc: 268.33 ng/ml




#28 AR-1254-3

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 234924
Conc: 141.19 ng/ml



#28 AR-1254-3


R.T.: 6.738 min
Delta R.T.: 0.014 min
Response: 457729
Conc: 304.27 ng/ml



#29 AR-1254-4


R.T.: 8.045 min
Delta R.T.: -0.004 min
Response: 123793
Conc: 101.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.960 min
Delta R.T.: -0.004 min
Response: 56721
Conc: 58.90 ng/ml




#30 AR-1254-5

R.T.: 8.476 min
Delta R.T.: -0.003 min
Response: 987983
Conc: 778.67 ng/ml



#30 AR-1254-5


R.T.: 7.391 min
Delta R.T.: -0.005 min
Response: 889816
Conc: 632.18 ng/ml



#31 AR-1260-1


R.T.: 7.918 min
Delta R.T.: 0.000 min
Response: 766288
Conc: 623.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.858 min
Delta R.T.: -0.003 min
Response: 629813
Conc: 584.83 ng/ml




#32 AR-1260-2

R.T.: 8.184 min
Delta R.T.: -0.001 min
Response: 877380
Conc: 633.75 ng/ml



#32 AR-1260-2


R.T.: 7.055 min
Delta R.T.: -0.003 min
Response: 759300
Conc: 592.83 ng/ml



#33 AR-1260-3


R.T.: 8.549 min
Delta R.T.: -0.002 min
Response: 493784
Conc: 597.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.212 min
Delta R.T.: -0.003 min
Response: 708260
Conc: 588.72 ng/ml




#34 AR-1260-4

R.T.: 8.781 min
Delta R.T.: -0.001 min
Response: 597929
Conc: 599.75 ng/ml



#34 AR-1260-4


R.T.: 7.697 min
Delta R.T.: -0.003 min
Response: 537151
Conc: 576.75 ng/ml



#35 AR-1260-5


R.T.: 9.106 min
Delta R.T.: -0.001 min
Response: 1073602
Conc: 566.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.945 min
Delta R.T.: -0.004 min
Response: 1201743
Conc: 560.27 ng/ml




#36 AR-1262-1

R.T.: 8.549 min
Delta R.T.: -0.002 min
Response: 493784
Conc: 369.80 ng/ml



#36 AR-1262-1


R.T.: 7.391 min
Delta R.T.: -0.003 min
Response: 889816
Conc: 1262.18 ng/ml



#37 AR-1262-2


R.T.: 9.106 min
Delta R.T.: -0.002 min
Response: 1073602
Conc: 488.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.945 min
Delta R.T.: -0.004 min
Response: 1201743
Conc: 529.00 ng/ml




#38 AR-1262-3

R.T.: 9.439 min
Delta R.T.: 0.021 min
Response: 726084
Conc: 743.79 ng/ml



#38 AR-1262-3


R.T.: 8.234 min
Delta R.T.: -0.004 min
Response: 253515
Conc: 249.43 ng/ml



#39 AR-1262-4


R.T.: 9.490 min
Delta R.T.: -0.021 min
Response: 410354
Conc: 686.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.297 min
Delta R.T.: -0.005 min
Response: 898205
Conc: 500.06 ng/ml




#40 AR-1262-5

R.T.: 10.176 min
Delta R.T.: -0.004 min
Response: 266608
Conc: 342.44 ng/ml



#40 AR-1262-5


R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 300658
Conc: 330.23 ng/ml



#41 AR-1268-1


R.T.: 9.439 min
Delta R.T.: 0.022 min
Response: 726084
Conc: 269.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.234 min
Delta R.T.: -0.005 min
Response: 253515
Conc: 86.94 ng/ml




#42 AR-1268-2

R.T.: 9.490 min
Delta R.T.: -0.024 min
Response: 410354
Conc: 159.31 ng/ml



#42 AR-1268-2


R.T.: 8.297 min
Delta R.T.: -0.006 min
Response: 898205
Conc: 336.95 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.740 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.509 min
Delta R.T.: -0.004 min
Response: 21287
Conc: 9.28 ng/ml




#44 AR-1268-4

R.T.: 10.176 min
Delta R.T.: -0.003 min
Response: 266608
Conc: 290.22 ng/ml



#44 AR-1268-4


R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 300658
Conc: 292.10 ng/ml



#45 AR-1268-5

R.T.: 10.591 min
Delta R.T.: -0.003 min
Response: 95294
Conc: 12.18 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.081 min
Delta R.T.: -0.006 min
Response: 99541
Conc: 13.57 ng/ml