

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093021\
 Data File : PP039697.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Sep 2021 23:04
 Operator : AJ\MA
 Sample : M3981-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 01:55:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.905	3.995	640998	379836	20.248	23.311
2)	SA Decachlor...	10.927	9.328	533561	538035	24.595	24.621
Target Compounds							
8)	L2 AR-1221-1	5.154	0.000	132059	0	351.871	N.D. #
9)	L2 AR-1221-2	0.000	4.349	0	10029	N.D.	58.850 #
27)	L6 AR-1254-2	0.000	6.308	0	8726	N.D.	9.054 #
28)	L6 AR-1254-3	7.767	6.725	35911	36238	21.583	24.089
30)	L6 AR-1254-5	8.460f	7.390	13039	25162	10.276	17.876 #
34)	L7 AR-1260-4	0.000	7.696	0	24928	N.D.	26.765 #
35)	L7 AR-1260-5	0.000	7.947	0	36270	N.D.	16.909 #
36)	L8 AR-1262-1	0.000	7.390	0	25162	N.D.	35.691 #
37)	L8 AR-1262-2	0.000	7.947	0	36270	N.D.	15.966 #
39)	L8 AR-1262-4	0.000	8.275f	0	52751	N.D.	29.368 #
42)	L9 AR-1268-2	0.000	8.275f	0	52751	N.D.	19.789 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093021\
Data File : PP039697.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Sep 2021 23:04
Operator : AJ\MA
Sample : M3981-02
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:55:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 25 01:28:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 640998
Conc: 20.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: 0.000 min
Response: 379836
Conc: 23.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.927 min
Delta R.T.: -0.002 min
Response: 533561
Conc: 24.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.328 min
Delta R.T.: -0.008 min
Response: 538035
Conc: 24.62 ng/ml



#8 AR-1221-1


R.T.: 5.154 min
Delta R.T.: -0.002 min
Response: 132059
Conc: 351.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.259 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.255 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.349 min
Delta R.T.: -0.009 min
Response: 10029
Conc: 58.85 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 7.371 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.308 min
Delta R.T.: 0.007 min
Response: 8726
Conc: 9.05 ng/ml




#28 AR-1254-3

R.T.: 7.767 min
Delta R.T.: 0.015 min
Response: 35911
Conc: 21.58 ng/ml



#28 AR-1254-3


R.T.: 6.725 min
Delta R.T.: 0.000 min
Response: 36238
Conc: 24.09 ng/ml



#30 AR-1254-5


R.T.: 8.460 min
Delta R.T.: -0.018 min
Response: 13039
Conc: 10.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 7.390 min
Delta R.T.: -0.006 min
Response: 25162
Conc: 17.88 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 8.782 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 7.696 min
Delta R.T.: -0.005 min
Response: 24928
Conc: 26.77 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 9.107 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.947 min
Delta R.T.: -0.002 min
Response: 36270
Conc: 16.91 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 8.552 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.390 min
Delta R.T.: -0.004 min
Response: 25162
Conc: 35.69 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 9.107 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.947 min
Delta R.T.: -0.002 min
Response: 36270
Conc: 15.97 ng/ml




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 9.511 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.275 min
Delta R.T.: -0.027 min
Response: 52751
Conc: 29.37 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 9.514 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#42 AR-1268-2

R.T.: 8.275 min
Delta R.T.: -0.028 min
Response: 52751
Conc: 19.79 ng/ml