

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093022\
 Data File : PP051811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2022 20:01
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 01:40:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 20 14:37:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.679	85078865	72146045	47.876	47.762
2) SA Decachlor...	10.247	8.774	65726869	56508724	42.529	47.161
Target Compounds						
16) L4 AR-1242-1	5.601	4.782	23817638	18564708	450.055	429.270
17) L4 AR-1242-2	5.623	4.801	33472607	25997865	448.260	430.590
18) L4 AR-1242-3	5.686	4.980	21277860	14060554	449.539	439.199
19) L4 AR-1242-4	5.785	5.065	16811261	14429989	457.495	443.793
20) L4 AR-1242-5	6.525	5.596	14311699	17632398	401.365	464.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093022\
Data File : PP051811.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2022 20:01
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:40:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091922.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 20 14:37:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.427 min
Delta R.T.: -0.005 min
Response: 85078865
Conc: 47.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.679 min
Delta R.T.: -0.006 min
Response: 72146045
Conc: 47.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.247 min
Delta R.T.: -0.007 min
Response: 65726869
Conc: 42.53 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.774 min
Delta R.T.: -0.014 min
Response: 56508724
Conc: 47.16 ng/ml



#16 AR-1242-1


R.T.: 5.601 min
Delta R.T.: -0.006 min
Response: 23817638
Conc: 450.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.782 min
Delta R.T.: -0.008 min
Response: 18564708
Conc: 429.27 ng/ml




#17 AR-1242-2

R.T.: 5.623 min
Delta R.T.: -0.005 min
Response: 33472607
Conc: 448.26 ng/ml



#17 AR-1242-2


R.T.: 4.801 min
Delta R.T.: -0.008 min
Response: 25997865
Conc: 430.59 ng/ml



#18 AR-1242-3


R.T.: 5.686 min
Delta R.T.: -0.006 min
Response: 21277860
Conc: 449.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.980 min
Delta R.T.: -0.008 min
Response: 14060554
Conc: 439.20 ng/ml




#19 AR-1242-4

R.T.: 5.785 min
Delta R.T.: -0.006 min
Response: 16811261
Conc: 457.49 ng/ml



#19 AR-1242-4


R.T.: 5.065 min
Delta R.T.: -0.009 min
Response: 14429989
Conc: 443.79 ng/ml



#20 AR-1242-5

R.T.: 6.525 min
Delta R.T.: -0.006 min
Response: 14311699
Conc: 401.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.596 min
Delta R.T.: -0.008 min
Response: 17632398
Conc: 464.31 ng/ml