

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
 Data File : PP075452.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2025 17:03
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 02:09:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.796	66706377	343.1E6	50.980	56.243
2) SA Decachlor...	10.432	8.799	52808458	351.6E6	52.938	58.288
Target Compounds						
26) L6 AR-1254-1	6.654	5.674	30136873	268.8E6	469.246m	538.986
27) L6 AR-1254-2	6.872	5.821	41871760	212.0E6	487.256	530.977
28) L6 AR-1254-3	7.233	6.223	44073998	350.6E6	476.876	527.429
29) L6 AR-1254-4	7.515	6.449	34384045	243.6E6	481.912	493.511
30) L6 AR-1254-5	7.931	6.865	40845969	282.3E6	499.295	494.741



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 17:03
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:09:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.001 min
Response: 66706377
Conc: 50.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 343129723
Conc: 56.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: -0.002 min
Response: 52808458
Conc: 52.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: -0.004 min
Response: 351619807
Conc: 58.29 ng/ml



#26 AR-1254-1


R.T.: 6.654 min
Delta R.T.: -0.005 min
Response: 30136873
Conc: 469.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.674 min
Delta R.T.: -0.004 min
Response: 268847637
Conc: 538.99 ng/ml




#27 AR-1254-2

R.T.: 6.872 min
Delta R.T.: -0.004 min
Response: 41871760
Conc: 487.26 ng/ml



#27 AR-1254-2


R.T.: 5.821 min
Delta R.T.: -0.005 min
Response: 211957619
Conc: 530.98 ng/ml



#28 AR-1254-3


R.T.: 7.233 min
Delta R.T.: -0.005 min
Response: 44073998
Conc: 476.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.223 min
Delta R.T.: -0.004 min
Response: 350594475
Conc: 527.43 ng/ml




#29 AR-1254-4

R.T.: 7.515 min
Delta R.T.: -0.004 min
Response: 34384045
Conc: 481.91 ng/ml



#29 AR-1254-4


R.T.: 6.449 min
Delta R.T.: -0.005 min
Response: 243620257
Conc: 493.51 ng/ml



#30 AR-1254-5

R.T.: 7.931 min
Delta R.T.: -0.005 min
Response: 40845969
Conc: 499.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.865 min
Delta R.T.: -0.006 min
Response: 282345268
Conc: 494.74 ng/ml