

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
 Data File : PP075464.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2025 21:07
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 02:19:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.797	64694786	330.7E6	49.443	54.202
2) SA Decachlor...	10.436	8.800	49484614	327.9E6	49.606	54.349
Target Compounds						
3) L1 AR-1016-1	5.808	4.891	22598438	298.8E6	469.775	532.579
4) L1 AR-1016-2	5.829	4.949	33637907	138.5E6	468.961	538.785
5) L1 AR-1016-3	5.892	5.069	21458883	76407074	461.176	492.097
6) L1 AR-1016-4	5.989	5.110	17645008	75727994	495.136	515.269
7) L1 AR-1016-5	6.281	5.324	16691593	82888707	494.987	507.330
31) L7 AR-1260-1	7.399	6.352	28866574	177.0E6	480.153	435.908
32) L7 AR-1260-2	7.652	6.539	33689387	254.6E6	462.779	442.142
33) L7 AR-1260-3	8.010	6.693	27013536	191.8E6	498.910	447.947
34) L7 AR-1260-4	8.238	7.162	29461102	177.6E6	521.093	411.900
35) L7 AR-1260-5	8.564	7.401	56539273	455.5E6	514.891	433.032



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075464.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 21:07
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:19:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 64694786
Conc: 49.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 330677315
Conc: 54.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.436 min
Delta R.T.: 0.001 min
Response: 49484614
Conc: 49.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.800 min
Delta R.T.: -0.004 min
Response: 327863044
Conc: 54.35 ng/ml



#3 AR-1016-1


R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 22598438
Conc: 469.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.891 min
Delta R.T.: -0.001 min
Response: 298783937
Conc: 532.58 ng/ml




#4 AR-1016-2

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 33637907
Conc: 468.96 ng/ml



#4 AR-1016-2


R.T.: 4.949 min
Delta R.T.: -0.002 min
Response: 138469010
Conc: 538.79 ng/ml



#5 AR-1016-3


R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 21458883
Conc: 461.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.069 min
Delta R.T.: -0.002 min
Response: 76407074
Conc: 492.10 ng/ml




#6 AR-1016-4

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 17645008
Conc: 495.14 ng/ml



#6 AR-1016-4


R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 75727994
Conc: 515.27 ng/ml



#7 AR-1016-5


R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 16691593
Conc: 494.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: -0.001 min
Response: 82888707
Conc: 507.33 ng/ml




#31 AR-1260-1

R.T.: 7.399 min
Delta R.T.: 0.000 min
Response: 28866574
Conc: 480.15 ng/ml



#31 AR-1260-1


R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 176950567
Conc: 435.91 ng/ml



#32 AR-1260-2


R.T.: 7.652 min
Delta R.T.: 0.000 min
Response: 33689387
Conc: 462.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.539 min
Delta R.T.: -0.002 min
Response: 254598789
Conc: 442.14 ng/ml




#33 AR-1260-3

R.T.: 8.010 min
Delta R.T.: 0.000 min
Response: 27013536
Conc: 498.91 ng/ml



#33 AR-1260-3


R.T.: 6.693 min
Delta R.T.: -0.003 min
Response: 191759366
Conc: 447.95 ng/ml



#34 AR-1260-4


R.T.: 8.238 min
Delta R.T.: 0.000 min
Response: 29461102
Conc: 521.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.162 min
Delta R.T.: -0.003 min
Response: 177621145
Conc: 411.90 ng/ml



#35 AR-1260-5

R.T.: 8.564 min
Delta R.T.: 0.000 min
Response: 56539273
Conc: 514.89 ng/ml



#35 AR-1260-5

R.T.: 7.401 min
Delta R.T.: -0.003 min
Response: 455544935
Conc: 433.03 ng/ml