

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100124\
 Data File : PP067395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2024 17:05
 Operator : YP\AJ
 Sample : P4224-03
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 1083

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/02/2024
 Supervised By :Ankita Jodhani 10/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 03:01:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.752	4.051	22587467	24606753	16.805m	16.469
2) SA Decachlor...	10.663	9.221	23775554	21535112	16.378	15.628

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100124\
Data File : PP067395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2024 17:05
Operator : YP\AJ
Sample : P4224-03
Misc :
ALS Vial : 40 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
1083


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 03:01:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.752 min
Delta R.T.: 0.000 min
Response: 22587467
Conc: 16.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
1083


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: -0.001 min
Response: 24606753
Conc: 16.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.663 min
Delta R.T.: 0.002 min
Response: 23775554
Conc: 16.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.221 min
Delta R.T.: -0.005 min
Response: 21535112
Conc: 15.63 ng/ml