

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100124\
 Data File : PP067419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2024 05:43
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/02/2024
 Supervised By :Ankita Jodhani 10/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 05:59:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.753	4.051	59297277	63615122	44.117	42.576
2) SA Decachlor...	10.662	9.218	63699891	61838734	43.879	44.875m
Target Compounds						
26) L6 AR-1254-1	6.772	5.975	28705160	35386112	460.732	436.536
27) L6 AR-1254-2	6.989	6.122	39052146	30975203	430.400	422.336
28) L6 AR-1254-3	7.354	6.533	38874229	45249165	414.444	396.856
29) L6 AR-1254-4	7.636	6.761	28017098	25592525	420.400	410.038m
30) L6 AR-1254-5	8.054	7.182	32284487	40288832	392.775m	392.413

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100124\
Data File : PP067419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2024 05:43
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 05:59:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.753 min
Delta R.T.: 0.000 min
Response: 59297277
Conc: 44.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: -0.002 min
Response: 63615122
Conc: 42.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.662 min
Delta R.T.: 0.000 min
Response: 63699891
Conc: 43.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.218 min
Delta R.T.: -0.008 min
Response: 61838734
Conc: 44.87 ng/ml m




#26 AR-1254-1

R.T.: 6.772 min
Delta R.T.: 0.002 min
Response: 28705160
Conc: 460.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024




#26 AR-1254-1

R.T.: 5.975 min
Delta R.T.: -0.005 min
Response: 35386112
Conc: 436.54 ng/ml




#27 AR-1254-2

R.T.: 6.989 min
Delta R.T.: 0.002 min
Response: 39052146
Conc: 430.40 ng/ml



#27 AR-1254-2

R.T.: 6.122 min
Delta R.T.: -0.005 min
Response: 30975203
Conc: 422.34 ng/ml




#28 AR-1254-3

R.T.: 7.354 min
Delta R.T.: 0.002 min
Response: 38874229
Conc: 414.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024




#28 AR-1254-3

R.T.: 6.533 min
Delta R.T.: -0.005 min
Response: 45249165
Conc: 396.86 ng/ml




#29 AR-1254-4

R.T.: 7.636 min
Delta R.T.: 0.002 min
Response: 28017098
Conc: 420.40 ng/ml



#29 AR-1254-4

R.T.: 6.761 min
Delta R.T.: -0.004 min
Response: 25592525
Conc: 410.04 ng/ml m




#30 AR-1254-5

R.T.: 8.054 min
Delta R.T.: 0.002 min
Response: 32284487
Conc: 392.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2024
Supervised By :Ankita Jodhani 10/02/2024



#30 AR-1254-5

R.T.: 7.182 min
Delta R.T.: -0.005 min
Response: 40288832
Conc: 392.41 ng/ml