

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100223\
 Data File : PP060513.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 20:25
 Operator : YP\AJ
 Sample : 04684-19
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 00:23:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	31094062	21441244	13.884	14.405
2) SA Decachlor...	10.457	8.747	26654820	22837933	15.097	14.001
Target Compounds						
3) L1 AR-1016-1	0.000	4.787	0	57300	N.D.	0.986 #
4) L1 AR-1016-2	5.755	4.805	1371481	72059	13.095	0.895 #
5) L1 AR-1016-3	0.000	4.980	0	50666	N.D.	1.134 #
6) L1 AR-1016-4	0.000	5.018	0	222940	N.D.	6.196 #
7) L1 AR-1016-5	6.202	5.237	833166	452544	14.979	9.314 #
8) L2 AR-1221-1	0.000	3.911	0	118852	N.D.	5.689 #
9) L2 AR-1221-2	4.846	3.989	606872	664485	28.405	44.564 #
10) L2 AR-1221-3	0.000	4.065	0	12169	N.D.	0.277 #
11) L3 AR-1232-1	0.000	4.065	0	12169	N.D.	0.334 #
12) L3 AR-1232-2	0.000	4.805	0	72059	N.D.	1.976 #
13) L3 AR-1232-3	5.755	4.980	1371481	50666	29.547	2.531 #
14) L3 AR-1232-4	5.940f	5.076	478813	293997	19.480	15.752
15) L3 AR-1232-5	6.018	5.237	411993	452544	19.882	21.758
16) L4 AR-1242-1	0.000	4.787	0	57300	N.D.	1.132 #
17) L4 AR-1242-2	5.755	4.805	1371481	72059	15.147	1.038 #
18) L4 AR-1242-3	0.000	4.980	0	50666	N.D.	1.316 #
19) L4 AR-1242-4	0.000	5.076	0	293997	N.D.	7.465 #
20) L4 AR-1242-5	6.668f	5.592	20049736	3018432	398.635	61.705 #
21) L5 AR-1248-1	0.000	4.781	0	64084	N.D.	1.756 #
22) L5 AR-1248-2	5.998	5.018	162734	222940	2.347	4.265 #
23) L5 AR-1248-3	6.202	5.076	833166	293997	10.970	5.295 #
24) L5 AR-1248-4	6.613	5.237	1955459	452544	23.644	6.875 #
25) L5 AR-1248-5	6.668f	5.642	20049736	1548761	249.118	24.078 #
26) L6 AR-1254-1	6.584	5.592	2971506	3018432	34.820	30.514
27) L6 AR-1254-2	6.807	5.740	8543224	4494979	66.939	52.302
28) L6 AR-1254-3	7.177	6.147	13248716	12074055	99.677	84.241
29) L6 AR-1254-4	7.465	6.377	11964911	9541021	121.670	109.462
30) L6 AR-1254-5	7.888	6.797	19428119	21078146	179.320	166.336
31) L7 AR-1260-1	7.344	6.278	7384292	7475492	70.444	74.475
32) L7 AR-1260-2	7.602	6.466	10016941	8482244	82.849	71.097
33) L7 AR-1260-3	7.951	6.621	3225722	8144641	34.801	71.136 #
34) L7 AR-1260-4	8.183	7.096	8041525	1518971	75.827	15.959 #
35) L7 AR-1260-5	8.530	7.338	2547611	2368323	13.165	10.928
36) L8 AR-1262-1	7.951	6.887	3225722	748425	25.861	13.977 #
37) L8 AR-1262-2	8.530	7.096	2547611	1518971	12.161	12.878
38) L8 AR-1262-3	8.875f	7.626	2766880	292233	18.283	3.169 #
39) L8 AR-1262-4	8.947	7.685	622613	1875077	5.202	11.131 #
40) L8 AR-1262-5	0.000	8.189	0	3157	N.D.	0.041 #
41) L9 AR-1268-1	8.875f	7.626	2766880	292233	10.429	1.075 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100223\
 Data File : PP060513.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 20:25
 Operator : YP\AJ
 Sample : 04684-19
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 00:23:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.947	7.685	622613	1875077	2.554	7.689 #
43) L9	AR-1268-3	0.000	7.897	0	157205	N.D.	0.735 #
44) L9	AR-1268-4	0.000	8.189	0	3157	N.D.	0.036 #
45) L9	AR-1268-5	0.000	8.487	0	280603	N.D.	0.449 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100223\
Data File : PP060513.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2023 20:25
Operator : YP\AJ
Sample : 04684-19
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 00:23:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 31094062
Conc: 13.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 21441244
Conc: 14.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.457 min
Delta R.T.: 0.002 min
Response: 26654820
Conc: 15.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: -0.004 min
Response: 22837933
Conc: 14.00 ng/ml



#3 AR-1016-1


R.T.: 0.000 min
Exp R.T.: 5.719 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.787 min
Delta R.T.: 0.002 min
Response: 57300
Conc: 0.99 ng/ml




#4 AR-1016-2

R.T.: 5.755 min
Delta R.T.: 0.013 min
Response: 1371481
Conc: 13.10 ng/ml



#4 AR-1016-2


R.T.: 4.805 min
Delta R.T.: 0.001 min
Response: 72059
Conc: 0.90 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T.: 5.806 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 50666
Conc: 1.13 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T.: 5.906 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.018 min
Delta R.T.: -0.006 min
Response: 222940
Conc: 6.20 ng/ml



#7 AR-1016-5


R.T.: 6.202 min
Delta R.T.: -0.002 min
Response: 833166
Conc: 14.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 452544
Conc: 9.31 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 4.743 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.911 min
Delta R.T.: 0.008 min
Response: 118852
Conc: 5.69 ng/ml



#9 AR-1221-2


R.T.: 4.846 min
Delta R.T.: 0.015 min
Response: 606872
Conc: 28.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.989 min
Delta R.T.: 0.000 min
Response: 664485
Conc: 44.56 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T.: 4.909 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.065 min
Delta R.T.: -0.001 min
Response: 12169
Conc: 0.28 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 4.926 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.065 min
Delta R.T.: -0.001 min
Response: 12169
Conc: 0.33 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T.: 5.464 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 72059
Conc: 1.98 ng/ml



#13 AR-1232-3


R.T.: 5.755 min
Delta R.T.: -0.004 min
Response: 1371481
Conc: 29.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 50666
Conc: 2.53 ng/ml




#14 AR-1232-4

R.T.: 5.940 min
Delta R.T.: 0.018 min
Response: 478813
Conc: 19.48 ng/ml



#14 AR-1232-4


R.T.: 5.076 min
Delta R.T.: 0.009 min
Response: 293997
Conc: 15.75 ng/ml



#15 AR-1232-5


R.T.: 6.018 min
Delta R.T.: 0.004 min
Response: 411993
Conc: 19.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.237 min
Delta R.T.: -0.003 min
Response: 452544
Conc: 21.76 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T.: 5.719 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 4.787 min
Delta R.T.: 0.003 min
Response: 57300
Conc: 1.13 ng/ml



#17 AR-1242-2


R.T.: 5.755 min
Delta R.T.: 0.013 min
Response: 1371481
Conc: 15.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.805 min
Delta R.T.: 0.001 min
Response: 72059
Conc: 1.04 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 5.805 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 4.980 min
Delta R.T.: -0.001 min
Response: 50666
Conc: 1.32 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 5.905 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.076 min
Delta R.T.: 0.010 min
Response: 293997
Conc: 7.47 ng/ml




#20 AR-1242-5

R.T.: 6.668 min
Delta R.T.: 0.017 min
Response: 20049736
Conc: 398.64 ng/ml



#20 AR-1242-5


R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 3018432
Conc: 61.70 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 5.719 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.781 min
Delta R.T.: -0.003 min
Response: 64084
Conc: 1.76 ng/ml




#22 AR-1248-2

R.T.: 5.998 min
Delta R.T.: 0.000 min
Response: 162734
Conc: 2.35 ng/ml



#22 AR-1248-2


R.T.: 5.018 min
Delta R.T.: -0.005 min
Response: 222940
Conc: 4.27 ng/ml



#23 AR-1248-3


R.T.: 6.202 min
Delta R.T.: -0.001 min
Response: 833166
Conc: 10.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.076 min
Delta R.T.: 0.010 min
Response: 293997
Conc: 5.29 ng/ml




#24 AR-1248-4

R.T.: 6.613 min
Delta R.T.: 0.000 min
Response: 1955459
Conc: 23.64 ng/ml



#24 AR-1248-4


R.T.: 5.237 min
Delta R.T.: -0.002 min
Response: 452544
Conc: 6.87 ng/ml



#25 AR-1248-5


R.T.: 6.668 min
Delta R.T.: 0.017 min
Response: 20049736
Conc: 249.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.642 min
Delta R.T.: 0.007 min
Response: 1548761
Conc: 24.08 ng/ml




#26 AR-1254-1

R.T.: 6.584 min
Delta R.T.: -0.002 min
Response: 2971506
Conc: 34.82 ng/ml



#26 AR-1254-1


R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 3018432
Conc: 30.51 ng/ml



#27 AR-1254-2


R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 8543224
Conc: 66.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.740 min
Delta R.T.: -0.002 min
Response: 4494979
Conc: 52.30 ng/ml




#28 AR-1254-3

R.T.: 7.177 min
Delta R.T.: 0.000 min
Response: 13248716
Conc: 99.68 ng/ml



#28 AR-1254-3


R.T.: 6.147 min
Delta R.T.: -0.002 min
Response: 12074055
Conc: 84.24 ng/ml



#29 AR-1254-4


R.T.: 7.465 min
Delta R.T.: 0.000 min
Response: 11964911
Conc: 121.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.377 min
Delta R.T.: -0.002 min
Response: 9541021
Conc: 109.46 ng/ml




#30 AR-1254-5

R.T.: 7.888 min
Delta R.T.: 0.001 min
Response: 19428119
Conc: 179.32 ng/ml



#30 AR-1254-5


R.T.: 6.797 min
Delta R.T.: -0.002 min
Response: 21078146
Conc: 166.34 ng/ml



#31 AR-1260-1


R.T.: 7.344 min
Delta R.T.: 0.000 min
Response: 7384292
Conc: 70.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.278 min
Delta R.T.: -0.003 min
Response: 7475492
Conc: 74.48 ng/ml




#32 AR-1260-2

R.T.: 7.602 min
Delta R.T.: 0.000 min
Response: 10016941
Conc: 82.85 ng/ml



#32 AR-1260-2


R.T.: 6.466 min
Delta R.T.: -0.003 min
Response: 8482244
Conc: 71.10 ng/ml



#33 AR-1260-3


R.T.: 7.951 min
Delta R.T.: -0.014 min
Response: 3225722
Conc: 34.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.621 min
Delta R.T.: -0.004 min
Response: 8144641
Conc: 71.14 ng/ml




#34 AR-1260-4

R.T.: 8.183 min
Delta R.T.: -0.014 min
Response: 8041525
Conc: 75.83 ng/ml



#34 AR-1260-4


R.T.: 7.096 min
Delta R.T.: -0.004 min
Response: 1518971
Conc: 15.96 ng/ml



#35 AR-1260-5


R.T.: 8.530 min
Delta R.T.: 0.000 min
Response: 2547611
Conc: 13.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.338 min
Delta R.T.: -0.003 min
Response: 2368323
Conc: 10.93 ng/ml




#36 AR-1262-1

R.T.: 7.951 min
Delta R.T.: -0.013 min
Response: 3225722
Conc: 25.86 ng/ml



#36 AR-1262-1


R.T.: 6.887 min
Delta R.T.: -0.004 min
Response: 748425
Conc: 13.98 ng/ml



#37 AR-1262-2


R.T.: 8.530 min
Delta R.T.: 0.000 min
Response: 2547611
Conc: 12.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.096 min
Delta R.T.: -0.002 min
Response: 1518971
Conc: 12.88 ng/ml




#38 AR-1262-3

R.T.: 8.875 min
Delta R.T.: 0.015 min
Response: 2766880
Conc: 18.28 ng/ml



#38 AR-1262-3


R.T.: 7.626 min
Delta R.T.: -0.001 min
Response: 292233
Conc: 3.17 ng/ml



#39 AR-1262-4


R.T.: 8.947 min
Delta R.T.: -0.005 min
Response: 622613
Conc: 5.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.685 min
Delta R.T.: -0.005 min
Response: 1875077
Conc: 11.13 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.646 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.189 min
Delta R.T.: 0.000 min
Response: 3157
Conc: 0.04 ng/ml



#41 AR-1268-1


R.T.: 8.875 min
Delta R.T.: 0.018 min
Response: 2766880
Conc: 10.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.626 min
Delta R.T.: -0.001 min
Response: 292233
Conc: 1.08 ng/ml




#42 AR-1268-2

R.T.: 8.947 min
Delta R.T.: -0.009 min
Response: 622613
Conc: 2.55 ng/ml



#42 AR-1268-2


R.T.: 7.685 min
Delta R.T.: -0.006 min
Response: 1875077
Conc: 7.69 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.198 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.897 min
Delta R.T.: -0.003 min
Response: 157205
Conc: 0.73 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.648 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.189 min
Delta R.T.: 0.000 min
Response: 3157
Conc: 0.04 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.091 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.487 min
Delta R.T.: 0.000 min
Response: 280603
Conc: 0.45 ng/ml