

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100223\
 Data File : PP060518.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 22:18
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 00:24:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	111.8E6	78619783	49.918	52.818
2) SA Decachlor...	10.455	8.746	88045502	79065318	49.869	48.471
Target Compounds						
26) L6 AR-1254-1	6.586	5.591	40884417	45330173	479.075	458.256
27) L6 AR-1254-2	6.807	5.739	60246279	39156976	472.052	455.616
28) L6 AR-1254-3	7.177	6.146	62927604	65213621	473.438	455.000
29) L6 AR-1254-4	7.465	6.376	46395815	39779541	471.794	456.381
30) L6 AR-1254-5	7.888	6.797	51237807	59251740	472.921	467.579



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100223\
Data File : PP060518.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2023 22:18
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 00:24:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 111792447
Conc: 49.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 78619783
Conc: 52.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.455 min
Delta R.T.: 0.000 min
Response: 88045502
Conc: 49.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.746 min
Delta R.T.: -0.005 min
Response: 79065318
Conc: 48.47 ng/ml



#26 AR-1254-1


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 40884417
Conc: 479.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.591 min
Delta R.T.: -0.003 min
Response: 45330173
Conc: 458.26 ng/ml




#27 AR-1254-2

R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 60246279
Conc: 472.05 ng/ml



#27 AR-1254-2


R.T.: 5.739 min
Delta R.T.: -0.002 min
Response: 39156976
Conc: 455.62 ng/ml



#28 AR-1254-3


R.T.: 7.177 min
Delta R.T.: 0.000 min
Response: 62927604
Conc: 473.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.146 min
Delta R.T.: -0.003 min
Response: 65213621
Conc: 455.00 ng/ml




#29 AR-1254-4

R.T.: 7.465 min
Delta R.T.: 0.001 min
Response: 46395815
Conc: 471.79 ng/ml



#29 AR-1254-4


R.T.: 6.376 min
Delta R.T.: -0.002 min
Response: 39779541
Conc: 456.38 ng/ml



#30 AR-1254-5

R.T.: 7.888 min
Delta R.T.: 0.000 min
Response: 51237807
Conc: 472.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: -0.002 min
Response: 59251740
Conc: 467.58 ng/ml