

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100318\
 Data File : PP027608.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Oct 2018 19:25
 Operator : EI\MA
 Sample : PB113476TB
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 06:49:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092418.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 25 01:51:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.045	12.158	91822600	131.7E6	362.056	304.290
Target Compounds							
1) T	Dalapon	3.272	2.945	51168793	54828801	85.667	100.808
6) T	MCCP	12.726	0.000	39122356	0	84.890	N.D. #
7) T	MCPA	0.000	12.982	0	54735637	N.D.	40.771 #
8) T	DICHLORPROP	0.000	13.594	0	16227043	N.D.	39.114 #
11) T	2,4,5-TP ...	15.237	0.000	6398713	0	2.192	N.D. #
12) T	2,4,5-T	15.586	15.897	10984504	23150677	4.395	11.699 #
13) T	2,4-DB	16.345	0.000	16566107	0	87.407	N.D. #
14) T	DINOSEB	17.831	0.000	9533469	0	4.952	N.D. #
15) T	Picloram	17.603	18.269	37953389	41756701	12.351	16.680 #
16) T	DCPA	0.000	18.143	0	14801982	N.D.	4.637 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100318\
Data File : PP027608.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Oct 2018 19:25
Operator : EI\MA
Sample : PB113476TB
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 06:49:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092418.M
Quant Title : 8080.M
QLast Update : Tue Sep 25 01:51:01 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.272 min
Delta R.T.: -0.040 min
Response: 51168793
Conc: 85.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.945 min
Delta R.T.: -0.038 min
Response: 54828801
Conc: 100.81 ng/ml




#4 2,4-DCAA

R.T.: 12.045 min
Delta R.T.: 0.000 min
Response: 91822600
Conc: 362.06 ng/ml



#4 2,4-DCAA


R.T.: 12.158 min
Delta R.T.: -0.002 min
Response: 131705307
Conc: 304.29 ng/ml



#6 MCPP


R.T.: 12.726 min
Delta R.T.: 0.022 min
Response: 39122356
Conc: 84.89 ug/ml

Instrument :
ECD_P
ClientSampleId :




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 12.656 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 12.948 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 12.982 min
Delta R.T.: -0.063 min
Response: 54735637
Conc: 40.77 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 13.563 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#8 DICHLORPROP

R.T.: 13.594 min
Delta R.T.: 0.010 min
Response: 16227043
Conc: 39.11 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 15.237 min
Delta R.T.: 0.044 min
Response: 6398713
Conc: 2.19 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 15.251 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 15.586 min
Delta R.T.: 0.002 min
Response: 10984504
Conc: 4.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#12 2,4,5-T

R.T.: 15.897 min
Delta R.T.: 0.004 min
Response: 23150677
Conc: 11.70 ng/ml




#13 2,4-DB

R.T.: 16.345 min
Delta R.T.: 0.000 min
Response: 16566107
Conc: 87.41 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T. : 16.498 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 17.831 min
Delta R.T.: -0.002 min
Response: 9533469
Conc: 4.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#14 DINOSEB

R.T.: 0.000 min
Exp R.T. : 16.885 min
Response: 0
Conc: N.D.




#15 Picloram

R.T.: 17.603 min
Delta R.T.: 0.006 min
Response: 37953389
Conc: 12.35 ng/ml



#15 Picloram


R.T.: 18.269 min
Delta R.T.: 0.003 min
Response: 41756701
Conc: 16.68 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 18.197 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#16 DCPA

R.T.: 18.143 min
Delta R.T.: -0.002 min
Response: 14801982
Conc: 4.64 ng/ml