

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100322\
 Data File : PP051832.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Oct 2022 17:47
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 01:26:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 01:26:27 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.423	3.674	186.9E6	138.0E6	97.898	98.388
2) SA Decachlor...	10.241	8.770	125.4E6	127.5E6	94.628	96.398
Target Compounds						
21) L5 AR-1248-1	5.597	4.778	28712274	27364272	952.574	948.387
22) L5 AR-1248-2	5.872	5.019	47292482	37502368	1023.862	948.625
23) L5 AR-1248-3	6.078	5.062	55577595	39709042	1047.851	950.067
24) L5 AR-1248-4	6.482	5.236	60886290	47338811	950.497	957.199
25) L5 AR-1248-5	6.521	5.636	59662161	47339589	951.984	965.181



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100322\
Data File : PP051832.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Oct 2022 17:47
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 01:26:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 01:26:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.423 min
Delta R.T.: 0.000 min
Response: 186908470
Conc: 97.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: 0.000 min
Response: 138035563
Conc: 98.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: 0.000 min
Response: 125433811
Conc: 94.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 127534229
Conc: 96.40 ng/ml



#21 AR-1248-1


R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 28712274
Conc: 952.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 27364272
Conc: 948.39 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: 0.000 min
Response: 47292482
Conc: 1023.86 ng/ml



#22 AR-1248-2


R.T.: 5.019 min
Delta R.T.: 0.000 min
Response: 37502368
Conc: 948.63 ng/ml



#23 AR-1248-3


R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 55577595
Conc: 1047.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.062 min
Delta R.T.: 0.000 min
Response: 39709042
Conc: 950.07 ng/ml




#24 AR-1248-4

R.T.: 6.482 min
Delta R.T.: 0.000 min
Response: 60886290
Conc: 950.50 ng/ml



#24 AR-1248-4


R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 47338811
Conc: 957.20 ng/ml



#25 AR-1248-5

R.T.: 6.521 min
Delta R.T.: 0.000 min
Response: 59662161
Conc: 951.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.636 min
Delta R.T.: 0.000 min
Response: 47339589
Conc: 965.18 ng/ml