

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100421\
 Data File : PP039774.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 18:00
 Operator : AJ\MA
 Sample : M4036-02MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 01:01:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.905	3.995	668839	366078	21.127	22.467
2)	SA Decachlor...	10.929	9.327	359034	348577	16.550	15.951
Target Compounds							
3)	L1 AR-1016-1	6.227	5.258	495510	333172	462.152	462.300
4)	L1 AR-1016-2	6.251	5.278	715040	455437	455.825	461.589
5)	L1 AR-1016-3	6.317	5.471	449237	255762	458.787	469.129
6)	L1 AR-1016-4	6.424	5.523	364464	200266	465.478	451.976
7)	L1 AR-1016-5	6.740	5.754	346628	258413	468.322	467.427
8)	L2 AR-1221-1	5.152	4.253	64678	34229	172.334	169.892
9)	L2 AR-1221-2	5.254	4.355	57045	51671	209.542	303.207 #
10)	L2 AR-1221-3	5.341	4.443	311465	189063	363.844	349.679
11)	L3 AR-1232-1	5.341	4.443	311465	189063	486.667	467.523
12)	L3 AR-1232-2	5.932	5.278	380053	455437	1055.728	1090.624
13)	L3 AR-1232-3	6.251	5.471	715040	255762	1098.911	1152.851
14)	L3 AR-1232-4	6.424	5.568	364464	251791	1087.310	1072.815
15)	L3 AR-1232-5	6.522	5.754	277404	258413	1203.640	1182.935
16)	L4 AR-1242-1	6.227	5.258	495510	333172	612.264	621.158
17)	L4 AR-1242-2	6.251	5.278	715040	455437	613.000	621.846
18)	L4 AR-1242-3	6.317	5.471	449237	255762	612.929	621.127
19)	L4 AR-1242-4	6.424	5.568	364464	251791	609.393	596.038
20)	L4 AR-1242-5	7.209	6.135	42988	202032	69.215	380.724 #
21)	L5 AR-1248-1	6.227	5.258	495510	333172	771.642	767.194
22)	L5 AR-1248-2	6.522	5.523	277404	200266	336.179	337.946
23)	L5 AR-1248-3	6.740	5.568	346628	251791	352.418	415.551
24)	L5 AR-1248-4	7.169	5.754	51234	258413	47.196	379.497 #
25)	L5 AR-1248-5	7.209	6.178	42988	29027	41.196	41.181
26)	L6 AR-1254-1	7.139	6.135	245300	202032	224.734	194.051
27)	L6 AR-1254-2	7.368	6.295	282226	195891	174.725	203.269
28)	L6 AR-1254-3	7.752	6.736	168485	356819	101.261	237.192 #
29)	L6 AR-1254-4	8.046	6.957	79786	44273	65.317	45.977 #
30)	L6 AR-1254-5	8.476	7.388	831276	673763	655.167	478.683 #
31)	L7 AR-1260-1	7.918	6.855	542454	480039	441.424	445.756
32)	L7 AR-1260-2	8.184	7.052	727421	549964	525.429	429.392
33)	L7 AR-1260-3	8.550	7.209	356492	515077	431.006	428.146
34)	L7 AR-1260-4	8.781	7.694	428126	374852	429.429	402.490
35)	L7 AR-1260-5	9.105	7.942	753882	828751	397.830	386.377
36)	L8 AR-1262-1	8.550	7.388	356492	673763	266.979	955.711 #
37)	L8 AR-1262-2	9.105	7.942	753882	828751	342.742	364.809
38)	L8 AR-1262-3	9.438f	8.230	517208	172078	529.823	169.304 #
39)	L8 AR-1262-4	9.491f	8.293	287074	627246	479.935	349.209 #
40)	L8 AR-1262-5	10.177	8.793	183906	208160	236.218	228.633
41)	L9 AR-1268-1	9.438f	8.230	517208	172078	192.063	59.014 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100421\
 Data File : PP039774.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 18:00
 Operator : AJ\MA
 Sample : M4036-02MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 01:01:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.491f	8.293	287074	627246	111.449	235.303 #
43) L9	AR-1268-3	0.000	8.505	0	13838	N.D.	6.031 #
44) L9	AR-1268-4	10.177	8.793	183906	208160	200.192	202.237
45) L9	AR-1268-5	10.592	9.078	71794	63931	9.174	8.712



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100421\
Data File : PP039774.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Oct 2021 18:00
Operator : AJ\MA
Sample : M4036-02MSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 01:01:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 25 01:28:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 668839
Conc: 21.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.001 min
Response: 366078
Conc: 22.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.929 min
Delta R.T.: 0.000 min
Response: 359034
Conc: 16.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.327 min
Delta R.T.: -0.009 min
Response: 348577
Conc: 15.95 ng/ml



#3 AR-1016-1


R.T.: 6.227 min
Delta R.T.: -0.001 min
Response: 495510
Conc: 462.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.258 min
Delta R.T.: -0.004 min
Response: 333172
Conc: 462.30 ng/ml




#4 AR-1016-2

R.T.: 6.251 min
Delta R.T.: -0.001 min
Response: 715040
Conc: 455.83 ng/ml



#4 AR-1016-2


R.T.: 5.278 min
Delta R.T.: -0.005 min
Response: 455437
Conc: 461.59 ng/ml



#5 AR-1016-3


R.T.: 6.317 min
Delta R.T.: -0.001 min
Response: 449237
Conc: 458.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.471 min
Delta R.T.: -0.005 min
Response: 255762
Conc: 469.13 ng/ml




#6 AR-1016-4

R.T.: 6.424 min
Delta R.T.: 0.000 min
Response: 364464
Conc: 465.48 ng/ml



#6 AR-1016-4


R.T.: 5.523 min
Delta R.T.: -0.004 min
Response: 200266
Conc: 451.98 ng/ml



#7 AR-1016-5


R.T.: 6.740 min
Delta R.T.: -0.001 min
Response: 346628
Conc: 468.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.754 min
Delta R.T.: -0.005 min
Response: 258413
Conc: 467.43 ng/ml




#8 AR-1221-1

R.T.: 5.152 min
Delta R.T.: -0.004 min
Response: 64678
Conc: 172.33 ng/ml



#8 AR-1221-1


R.T.: 4.253 min
Delta R.T.: -0.005 min
Response: 34229
Conc: 169.89 ng/ml



#9 AR-1221-2


R.T.: 5.254 min
Delta R.T.: -0.001 min
Response: 57045
Conc: 209.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.355 min
Delta R.T.: -0.003 min
Response: 51671
Conc: 303.21 ng/ml




#10 AR-1221-3

R.T.: 5.341 min
Delta R.T.: -0.001 min
Response: 311465
Conc: 363.84 ng/ml



#10 AR-1221-3


R.T.: 4.443 min
Delta R.T.: -0.004 min
Response: 189063
Conc: 349.68 ng/ml



#11 AR-1232-1


R.T.: 5.341 min
Delta R.T.: 0.000 min
Response: 311465
Conc: 486.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.443 min
Delta R.T.: -0.003 min
Response: 189063
Conc: 467.52 ng/ml




#12 AR-1232-2

R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 380053
Conc: 1055.73 ng/ml



#12 AR-1232-2


R.T.: 5.278 min
Delta R.T.: -0.004 min
Response: 455437
Conc: 1090.62 ng/ml



#13 AR-1232-3


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 715040
Conc: 1098.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.471 min
Delta R.T.: -0.004 min
Response: 255762
Conc: 1152.85 ng/ml




#14 AR-1232-4

R.T.: 6.424 min
Delta R.T.: -0.001 min
Response: 364464
Conc: 1087.31 ng/ml



#14 AR-1232-4


R.T.: 5.568 min
Delta R.T.: -0.005 min
Response: 251791
Conc: 1072.81 ng/ml



#15 AR-1232-5


R.T.: 6.522 min
Delta R.T.: -0.001 min
Response: 277404
Conc: 1203.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.754 min
Delta R.T.: -0.004 min
Response: 258413
Conc: 1182.93 ng/ml




#16 AR-1242-1

R.T.: 6.227 min
Delta R.T.: -0.001 min
Response: 495510
Conc: 612.26 ng/ml



#16 AR-1242-1


R.T.: 5.258 min
Delta R.T.: -0.004 min
Response: 333172
Conc: 621.16 ng/ml



#17 AR-1242-2


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 715040
Conc: 613.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.278 min
Delta R.T.: -0.005 min
Response: 455437
Conc: 621.85 ng/ml




#18 AR-1242-3

R.T.: 6.317 min
Delta R.T.: -0.001 min
Response: 449237
Conc: 612.93 ng/ml



#18 AR-1242-3


R.T.: 5.471 min
Delta R.T.: -0.005 min
Response: 255762
Conc: 621.13 ng/ml



#19 AR-1242-4


R.T.: 6.424 min
Delta R.T.: -0.002 min
Response: 364464
Conc: 609.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.568 min
Delta R.T.: -0.005 min
Response: 251791
Conc: 596.04 ng/ml




#20 AR-1242-5

R.T.: 7.209 min
Delta R.T.: -0.002 min
Response: 42988
Conc: 69.22 ng/ml



#20 AR-1242-5


R.T.: 6.135 min
Delta R.T.: -0.004 min
Response: 202032
Conc: 380.72 ng/ml



#21 AR-1248-1


R.T.: 6.227 min
Delta R.T.: -0.001 min
Response: 495510
Conc: 771.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.258 min
Delta R.T.: -0.005 min
Response: 333172
Conc: 767.19 ng/ml




#22 AR-1248-2

R.T.: 6.522 min
Delta R.T.: -0.001 min
Response: 277404
Conc: 336.18 ng/ml



#22 AR-1248-2


R.T.: 5.523 min
Delta R.T.: -0.005 min
Response: 200266
Conc: 337.95 ng/ml



#23 AR-1248-3


R.T.: 6.740 min
Delta R.T.: -0.002 min
Response: 346628
Conc: 352.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.568 min
Delta R.T.: -0.005 min
Response: 251791
Conc: 415.55 ng/ml




#24 AR-1248-4

R.T.: 7.169 min
Delta R.T.: -0.002 min
Response: 51234
Conc: 47.20 ng/ml



#24 AR-1248-4


R.T.: 5.754 min
Delta R.T.: -0.004 min
Response: 258413
Conc: 379.50 ng/ml



#25 AR-1248-5


R.T.: 7.209 min
Delta R.T.: -0.001 min
Response: 42988
Conc: 41.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.178 min
Delta R.T.: -0.005 min
Response: 29027
Conc: 41.18 ng/ml




#26 AR-1254-1

R.T.: 7.139 min
Delta R.T.: -0.002 min
Response: 245300
Conc: 224.73 ng/ml



#26 AR-1254-1


R.T.: 6.135 min
Delta R.T.: -0.005 min
Response: 202032
Conc: 194.05 ng/ml



#27 AR-1254-2


R.T.: 7.368 min
Delta R.T.: -0.003 min
Response: 282226
Conc: 174.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.295 min
Delta R.T.: -0.006 min
Response: 195891
Conc: 203.27 ng/ml




#28 AR-1254-3

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 168485
Conc: 101.26 ng/ml



#28 AR-1254-3


R.T.: 6.736 min
Delta R.T.: 0.012 min
Response: 356819
Conc: 237.19 ng/ml



#29 AR-1254-4


R.T.: 8.046 min
Delta R.T.: -0.004 min
Response: 79786
Conc: 65.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.957 min
Delta R.T.: -0.006 min
Response: 44273
Conc: 45.98 ng/ml




#30 AR-1254-5

R.T.: 8.476 min
Delta R.T.: -0.002 min
Response: 831276
Conc: 655.17 ng/ml



#30 AR-1254-5


R.T.: 7.388 min
Delta R.T.: -0.008 min
Response: 673763
Conc: 478.68 ng/ml



#31 AR-1260-1


R.T.: 7.918 min
Delta R.T.: 0.000 min
Response: 542454
Conc: 441.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.855 min
Delta R.T.: -0.006 min
Response: 480039
Conc: 445.76 ng/ml




#32 AR-1260-2

R.T.: 8.184 min
Delta R.T.: -0.002 min
Response: 727421
Conc: 525.43 ng/ml



#32 AR-1260-2


R.T.: 7.052 min
Delta R.T.: -0.005 min
Response: 549964
Conc: 429.39 ng/ml



#33 AR-1260-3


R.T.: 8.550 min
Delta R.T.: -0.001 min
Response: 356492
Conc: 431.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.209 min
Delta R.T.: -0.006 min
Response: 515077
Conc: 428.15 ng/ml




#34 AR-1260-4

R.T.: 8.781 min
Delta R.T.: -0.001 min
Response: 428126
Conc: 429.43 ng/ml



#34 AR-1260-4


R.T.: 7.694 min
Delta R.T.: -0.007 min
Response: 374852
Conc: 402.49 ng/ml



#35 AR-1260-5


R.T.: 9.105 min
Delta R.T.: -0.002 min
Response: 753882
Conc: 397.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.942 min
Delta R.T.: -0.007 min
Response: 828751
Conc: 386.38 ng/ml




#36 AR-1262-1

R.T.: 8.550 min
Delta R.T.: -0.001 min
Response: 356492
Conc: 266.98 ng/ml



#36 AR-1262-1


R.T.: 7.388 min
Delta R.T.: -0.006 min
Response: 673763
Conc: 955.71 ng/ml



#37 AR-1262-2


R.T.: 9.105 min
Delta R.T.: -0.002 min
Response: 753882
Conc: 342.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.942 min
Delta R.T.: -0.007 min
Response: 828751
Conc: 364.81 ng/ml




#38 AR-1262-3

R.T.: 9.438 min
Delta R.T.: 0.021 min
Response: 517208
Conc: 529.82 ng/ml



#38 AR-1262-3


R.T.: 8.230 min
Delta R.T.: -0.008 min
Response: 172078
Conc: 169.30 ng/ml



#39 AR-1262-4


R.T.: 9.491 min
Delta R.T.: -0.020 min
Response: 287074
Conc: 479.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.293 min
Delta R.T.: -0.009 min
Response: 627246
Conc: 349.21 ng/ml




#40 AR-1262-5

R.T.: 10.177 min
Delta R.T.: -0.003 min
Response: 183906
Conc: 236.22 ng/ml



#40 AR-1262-5


R.T.: 8.793 min
Delta R.T.: -0.008 min
Response: 208160
Conc: 228.63 ng/ml



#41 AR-1268-1


R.T.: 9.438 min
Delta R.T.: 0.021 min
Response: 517208
Conc: 192.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.230 min
Delta R.T.: -0.009 min
Response: 172078
Conc: 59.01 ng/ml




#42 AR-1268-2

R.T.: 9.491 min
Delta R.T.: -0.022 min
Response: 287074
Conc: 111.45 ng/ml



#42 AR-1268-2


R.T.: 8.293 min
Delta R.T.: -0.010 min
Response: 627246
Conc: 235.30 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.740 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.505 min
Delta R.T.: -0.008 min
Response: 13838
Conc: 6.03 ng/ml




#44 AR-1268-4

R.T.: 10.177 min
Delta R.T.: -0.002 min
Response: 183906
Conc: 200.19 ng/ml



#44 AR-1268-4


R.T.: 8.793 min
Delta R.T.: -0.007 min
Response: 208160
Conc: 202.24 ng/ml



#45 AR-1268-5

R.T.: 10.592 min
Delta R.T.: -0.002 min
Response: 71794
Conc: 9.17 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.078 min
Delta R.T.: -0.009 min
Response: 63931
Conc: 8.71 ng/ml