

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100421\
 Data File : PP039761.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Oct 2021 14:22
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 16:38:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 25 01:28:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.906	3.995	1552605	883510	49.044	54.222
2) SA Decachlor...	10.929	9.327	1140916	1123660	52.592	51.421
Target Compounds						
26) L6 AR-1254-1	7.139	6.135	544118	507729	498.499	487.671
27) L6 AR-1254-2	7.369	6.295	812762	467859	503.177	485.480
28) L6 AR-1254-3	7.751	6.718	866059	732332	520.512	486.810
29) L6 AR-1254-4	8.048	6.958	624426	463431	511.182	481.268
30) L6 AR-1254-5	8.477	7.390	648141	701781	510.829	498.589



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100421\
Data File : PP039761.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Oct 2021 14:22
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 16:38:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092421.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 25 01:28:12 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 1552605
Conc: 49.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.001 min
Response: 883510
Conc: 54.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.929 min
Delta R.T.: 0.000 min
Response: 1140916
Conc: 52.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.327 min
Delta R.T.: -0.009 min
Response: 1123660
Conc: 51.42 ng/ml



#26 AR-1254-1


R.T.: 7.139 min
Delta R.T.: -0.001 min
Response: 544118
Conc: 498.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.135 min
Delta R.T.: -0.005 min
Response: 507729
Conc: 487.67 ng/ml




#27 AR-1254-2

R.T.: 7.369 min
Delta R.T.: -0.001 min
Response: 812762
Conc: 503.18 ng/ml



#27 AR-1254-2


R.T.: 6.295 min
Delta R.T.: -0.006 min
Response: 467859
Conc: 485.48 ng/ml



#28 AR-1254-3


R.T.: 7.751 min
Delta R.T.: -0.002 min
Response: 866059
Conc: 520.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.718 min
Delta R.T.: -0.006 min
Response: 732332
Conc: 486.81 ng/ml




#29 AR-1254-4

R.T.: 8.048 min
Delta R.T.: -0.002 min
Response: 624426
Conc: 511.18 ng/ml



#29 AR-1254-4


R.T.: 6.958 min
Delta R.T.: -0.006 min
Response: 463431
Conc: 481.27 ng/ml



#30 AR-1254-5

R.T.: 8.477 min
Delta R.T.: -0.002 min
Response: 648141
Conc: 510.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.390 min
Delta R.T.: -0.006 min
Response: 701781
Conc: 498.59 ng/ml