

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100423\
 Data File : PP060572.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Oct 2023 13:45
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/05/2023
 Supervised By :Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 15:30:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.533	3.686	120.2E6	84108227	53.670	56.505
2) SA Decachlor...	10.458	8.748	87081653	71061879	49.323	43.564
Target Compounds						
26) L6 AR-1254-1	6.587	5.591	44164573	49242947	517.511	497.812
27) L6 AR-1254-2	6.808	5.740	65610420	41470676	514.082	482.537
28) L6 AR-1254-3	7.178	6.147	66957034	67194094	503.753	468.818
29) L6 AR-1254-4	7.466	6.376	47957038	40231038	487.670	461.561
30) L6 AR-1254-5	7.888	6.797	50694976	57439152	467.911m	453.275

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100423\
Data File : PP060572.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Oct 2023 13:45
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 15:30:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.533 min
Delta R.T.: 0.001 min
Response: 120192916
Conc: 53.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023




#1 Tetrachloro-m-xylene

R.T.: 3.686 min
Delta R.T.: 0.000 min
Response: 84108227
Conc: 56.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.458 min
Delta R.T.: 0.002 min
Response: 87081653
Conc: 49.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.748 min
Delta R.T.: -0.002 min
Response: 71061879
Conc: 43.56 ng/ml




#26 AR-1254-1

R.T.: 6.587 min
Delta R.T.: 0.000 min
Response: 44164573
Conc: 517.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023




#26 AR-1254-1

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 49242947
Conc: 497.81 ng/ml




#27 AR-1254-2

R.T.: 6.808 min
Delta R.T.: 0.001 min
Response: 65610420
Conc: 514.08 ng/ml



#27 AR-1254-2

R.T.: 5.740 min
Delta R.T.: -0.002 min
Response: 41470676
Conc: 482.54 ng/ml




#28 AR-1254-3

R.T.: 7.178 min
Delta R.T.: 0.001 min
Response: 66957034
Conc: 503.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023




#28 AR-1254-3

R.T.: 6.147 min
Delta R.T.: -0.003 min
Response: 67194094
Conc: 468.82 ng/ml




#29 AR-1254-4

R.T.: 7.466 min
Delta R.T.: 0.002 min
Response: 47957038
Conc: 487.67 ng/ml



#29 AR-1254-4

R.T.: 6.376 min
Delta R.T.: -0.002 min
Response: 40231038
Conc: 461.56 ng/ml




#30 AR-1254-5

R.T.: 7.888 min
Delta R.T.: 0.001 min
Response: 50694976
Conc: 467.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/05/2023
Supervised By :Ankita Jodhani 10/05/2023



#30 AR-1254-5

R.T.: 6.797 min
Delta R.T.: -0.001 min
Response: 57439152
Conc: 453.28 ng/ml