

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
 Data File : PP051980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 15:13
 Operator : YP\AJ
 Sample : N4968-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CLI-0120

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 02:05:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.426	3.680	23093659	18804864	12.921	14.791
2)	SA Decachlor...	10.250	8.776	17188592	10452185	14.178	8.849 #
Target Compounds							
3)	L1 AR-1016-1	5.600	4.800f	5321609	47189025	117.421	1076.103 #
4)	L1 AR-1016-2	5.625	4.800	4600648	47189025	71.681	764.880 #
5)	L1 AR-1016-3	5.685	4.983	8291934	16650900	207.431	506.758 #
6)	L1 AR-1016-4	5.779	5.034f	10277715	20847533	293.167	806.091 #
7)	L1 AR-1016-5	6.072	5.245	6284745	2765036	152.180	80.290 #
8)	L2 AR-1221-1	4.637	3.895	21054638	2196893	976.714	140.269 #
9)	L2 AR-1221-2	4.713	3.981	4955230	3837670	310.858	321.217
10)	L2 AR-1221-3	4.798	4.063	4248712	3614677	87.742	96.349
11)	L3 AR-1232-1	4.798	4.063	4248712	3614677	105.883	116.118
12)	L3 AR-1232-2	5.333	4.800	3985897	47189025	212.396	1669.824 #
13)	L3 AR-1232-3	5.625	4.983	4600648	16650900	154.923	1140.197 #
14)	L3 AR-1232-4	5.779	5.034f	10277715	20847533	718.018	1562.189 #
15)	L3 AR-1232-5	5.879	5.245	7705791	2765036	644.070	187.227 #
16)	L4 AR-1242-1	5.600	4.800f	5321609	47189025	129.807	1210.990 #
17)	L4 AR-1242-2	5.625	4.800	4600648	47189025	79.999	861.105 #
18)	L4 AR-1242-3	5.685	4.983	8291934	16650900	231.575	565.377 #
19)	L4 AR-1242-4	5.779	5.034f	10277715	20847533	356.938	706.772 #
20)	L4 AR-1242-5	6.518	5.582	13437549	1657491	356.774	45.403 #
21)	L5 AR-1248-1	5.600	4.800f	5321609	47189025	170.166	1562.609 #
22)	L5 AR-1248-2	5.879	5.034f	7705791	20847533	145.986	501.234 #
23)	L5 AR-1248-3	6.072	5.034f	6284745	20847533	101.599	445.564 #
24)	L5 AR-1248-4	6.477	5.245	10565118	2765036	153.257	53.940 #
25)	L5 AR-1248-5	6.518	5.638	13437549	9218852	196.892	173.487
26)	L6 AR-1254-1	6.437f	5.582	2424241	1657491	35.528	22.280 #
27)	L6 AR-1254-2	6.692f	5.763f	20370566	4665531	199.057	72.350 #
28)	L6 AR-1254-3	7.035	0.000	13886450	0	135.630	N.D. #
29)	L6 AR-1254-4	7.352f	6.379	18037965	3566814	251.449	52.326 #
30)	L6 AR-1254-5	7.752	6.818	3099456	3665232	41.199	39.618
31)	L7 AR-1260-1	7.202	6.275	5295765	2616376	74.503	37.985 #
32)	L7 AR-1260-2	7.481	6.484	23741156	1186929	290.674	14.104 #
33)	L7 AR-1260-3	7.813f	6.654f	11857957	5286581	207.328	67.913 #
34)	L7 AR-1260-4	8.066	7.101	16272641	5432732	248.458	86.300 #
35)	L7 AR-1260-5	8.373	7.321f	2335910	14420772	18.371	98.326 #
36)	L8 AR-1262-1	7.813	6.881f	11857957	7005137	136.367	186.397 #
37)	L8 AR-1262-2	8.373	7.101	2335910	5432732	15.832	64.142 #
38)	L8 AR-1262-3	8.678f	7.649	1724863	625351	16.754	8.917 #
39)	L8 AR-1262-4	8.789	7.723f	12322615	11553474	267.660	92.468 #
40)	L8 AR-1262-5	0.000	8.215	0	7639322	N.D.	134.833 #
41)	L9 AR-1268-1	8.678f	7.649	1724863	625351	9.670	3.150 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
 Data File : PP051980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 15:13
 Operator : YP\AJ
 Sample : N4968-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CLI-0120

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 02:05:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.789	7.723f	12322615	11553474	74.333	64.651
43) L9	AR-1268-3	9.025	7.906	358108	497544	2.498	3.253 #
44) L9	AR-1268-4	0.000	8.215	0	7639322	N.D.	121.706 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
Data File : PP051980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 15:13
Operator : YP\AJ
Sample : N4968-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CLI-0120

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 02:05:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: 0.003 min
Response: 23093659
Conc: 12.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: 0.006 min
Response: 18804864
Conc: 14.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: 0.010 min
Response: 17188592
Conc: 14.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.776 min
Delta R.T.: 0.006 min
Response: 10452185
Conc: 8.85 ng/ml



#3 AR-1016-1


R.T.: 5.600 min
Delta R.T.: 0.004 min
Response: 5321609
Conc: 117.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#3 AR-1016-1

R.T.: 4.800 min
Delta R.T.: 0.022 min
Response: 47189025
Conc: 1076.10 ng/ml




#4 AR-1016-2

R.T.: 5.625 min
Delta R.T.: 0.006 min
Response: 4600648
Conc: 71.68 ng/ml



#4 AR-1016-2


R.T.: 4.800 min
Delta R.T.: 0.003 min
Response: 47189025
Conc: 764.88 ng/ml



#5 AR-1016-3


R.T.: 5.685 min
Delta R.T.: 0.003 min
Response: 8291934
Conc: 207.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#5 AR-1016-3

R.T.: 4.983 min
Delta R.T.: 0.007 min
Response: 16650900
Conc: 506.76 ng/ml




#6 AR-1016-4

R.T.: 5.779 min
Delta R.T.: -0.002 min
Response: 10277715
Conc: 293.17 ng/ml



#6 AR-1016-4


R.T.: 5.034 min
Delta R.T.: 0.015 min
Response: 20847533
Conc: 806.09 ng/ml



#7 AR-1016-5


R.T.: 6.072 min
Delta R.T.: -0.006 min
Response: 6284745
Conc: 152.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#7 AR-1016-5

R.T.: 5.245 min
Delta R.T.: 0.009 min
Response: 2765036
Conc: 80.29 ng/ml




#8 AR-1221-1

R.T.: 4.637 min
Delta R.T.: 0.006 min
Response: 21054638
Conc: 976.71 ng/ml



#8 AR-1221-1


R.T.: 3.895 min
Delta R.T.: 0.003 min
Response: 2196893
Conc: 140.27 ng/ml



#9 AR-1221-2


R.T.: 4.713 min
Delta R.T.: -0.005 min
Response: 4955230
Conc: 310.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#9 AR-1221-2

R.T.: 3.981 min
Delta R.T.: 0.002 min
Response: 3837670
Conc: 321.22 ng/ml




#10 AR-1221-3

R.T.: 4.798 min
Delta R.T.: 0.003 min
Response: 4248712
Conc: 87.74 ng/ml



#10 AR-1221-3


R.T.: 4.063 min
Delta R.T.: 0.007 min
Response: 3614677
Conc: 96.35 ng/ml



#11 AR-1232-1


R.T.: 4.798 min
Delta R.T.: 0.002 min
Response: 4248712
Conc: 105.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#11 AR-1232-1

R.T.: 4.063 min
Delta R.T.: 0.007 min
Response: 3614677
Conc: 116.12 ng/ml




#12 AR-1232-2

R.T.: 5.333 min
Delta R.T.: 0.004 min
Response: 3985897
Conc: 212.40 ng/ml



#12 AR-1232-2


R.T.: 4.800 min
Delta R.T.: 0.003 min
Response: 47189025
Conc: 1669.82 ng/ml



#13 AR-1232-3


R.T.: 5.625 min
Delta R.T.: 0.006 min
Response: 4600648
Conc: 154.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#13 AR-1232-3

R.T.: 4.983 min
Delta R.T.: 0.007 min
Response: 16650900
Conc: 1140.20 ng/ml




#14 AR-1232-4

R.T.: 5.779 min
Delta R.T.: -0.004 min
Response: 10277715
Conc: 718.02 ng/ml



#14 AR-1232-4


R.T.: 5.034 min
Delta R.T.: -0.028 min
Response: 20847533
Conc: 1562.19 ng/ml



#15 AR-1232-5


R.T.: 5.879 min
Delta R.T.: 0.006 min
Response: 7705791
Conc: 644.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#15 AR-1232-5

R.T.: 5.245 min
Delta R.T.: 0.009 min
Response: 2765036
Conc: 187.23 ng/ml




#16 AR-1242-1

R.T.: 5.600 min
Delta R.T.: 0.004 min
Response: 5321609
Conc: 129.81 ng/ml



#16 AR-1242-1


R.T.: 4.800 min
Delta R.T.: 0.022 min
Response: 47189025
Conc: 1210.99 ng/ml



#17 AR-1242-2


R.T.: 5.625 min
Delta R.T.: 0.007 min
Response: 4600648
Conc: 80.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#17 AR-1242-2

R.T.: 4.800 min
Delta R.T.: 0.003 min
Response: 47189025
Conc: 861.11 ng/ml




#18 AR-1242-3

R.T.: 5.685 min
Delta R.T.: 0.003 min
Response: 8291934
Conc: 231.58 ng/ml



#18 AR-1242-3


R.T.: 4.983 min
Delta R.T.: 0.006 min
Response: 16650900
Conc: 565.38 ng/ml



#19 AR-1242-4


R.T.: 5.779 min
Delta R.T.: -0.002 min
Response: 10277715
Conc: 356.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#19 AR-1242-4

R.T.: 5.034 min
Delta R.T.: -0.028 min
Response: 20847533
Conc: 706.77 ng/ml




#20 AR-1242-5

R.T.: 6.518 min
Delta R.T.: -0.004 min
Response: 13437549
Conc: 356.77 ng/ml



#20 AR-1242-5


R.T.: 5.582 min
Delta R.T.: -0.011 min
Response: 1657491
Conc: 45.40 ng/ml



#21 AR-1248-1


R.T.: 5.600 min
Delta R.T.: 0.004 min
Response: 5321609
Conc: 170.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#21 AR-1248-1

R.T.: 4.800 min
Delta R.T.: 0.022 min
Response: 47189025
Conc: 1562.61 ng/ml




#22 AR-1248-2

R.T.: 5.879 min
Delta R.T.: 0.007 min
Response: 7705791
Conc: 145.99 ng/ml



#22 AR-1248-2


R.T.: 5.034 min
Delta R.T.: 0.015 min
Response: 20847533
Conc: 501.23 ng/ml



#23 AR-1248-3


R.T.: 6.072 min
Delta R.T.: -0.005 min
Response: 6284745
Conc: 101.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#23 AR-1248-3

R.T.: 5.034 min
Delta R.T.: -0.028 min
Response: 20847533
Conc: 445.56 ng/ml




#24 AR-1248-4

R.T.: 6.477 min
Delta R.T.: -0.006 min
Response: 10565118
Conc: 153.26 ng/ml



#24 AR-1248-4


R.T.: 5.245 min
Delta R.T.: 0.009 min
Response: 2765036
Conc: 53.94 ng/ml



#25 AR-1248-5


R.T.: 6.518 min
Delta R.T.: -0.004 min
Response: 13437549
Conc: 196.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#25 AR-1248-5

R.T.: 5.638 min
Delta R.T.: 0.002 min
Response: 9218852
Conc: 173.49 ng/ml




#26 AR-1254-1

R.T.: 6.437 min
Delta R.T.: -0.020 min
Response: 2424241
Conc: 35.53 ng/ml



#26 AR-1254-1


R.T.: 5.582 min
Delta R.T.: -0.011 min
Response: 1657491
Conc: 22.28 ng/ml



#27 AR-1254-2


R.T.: 6.692 min
Delta R.T.: 0.016 min
Response: 20370566
Conc: 199.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#27 AR-1254-2

R.T.: 5.763 min
Delta R.T.: 0.020 min
Response: 4665531
Conc: 72.35 ng/ml




#28 AR-1254-3

R.T.: 7.035 min
Delta R.T.: -0.009 min
Response: 13886450
Conc: 135.63 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.153 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.352 min
Delta R.T.: 0.022 min
Response: 18037965
Conc: 251.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#29 AR-1254-4

R.T.: 6.379 min
Delta R.T.: -0.004 min
Response: 3566814
Conc: 52.33 ng/ml




#30 AR-1254-5

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 3099456
Conc: 41.20 ng/ml



#30 AR-1254-5


R.T.: 6.818 min
Delta R.T.: 0.012 min
Response: 3665232
Conc: 39.62 ng/ml



#31 AR-1260-1


R.T.: 7.202 min
Delta R.T.: -0.008 min
Response: 5295765
Conc: 74.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#31 AR-1260-1

R.T.: 6.275 min
Delta R.T.: -0.011 min
Response: 2616376
Conc: 37.98 ng/ml




#32 AR-1260-2

R.T.: 7.481 min
Delta R.T.: 0.014 min
Response: 23741156
Conc: 290.67 ng/ml



#32 AR-1260-2


R.T.: 6.484 min
Delta R.T.: 0.009 min
Response: 1186929
Conc: 14.10 ng/ml



#33 AR-1260-3


R.T.: 7.813 min
Delta R.T.: -0.016 min
Response: 11857957
Conc: 207.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#33 AR-1260-3

R.T.: 6.654 min
Delta R.T.: 0.023 min
Response: 5286581
Conc: 67.91 ng/ml




#34 AR-1260-4

R.T.: 8.066 min
Delta R.T.: 0.012 min
Response: 16272641
Conc: 248.46 ng/ml



#34 AR-1260-4


R.T.: 7.101 min
Delta R.T.: -0.007 min
Response: 5432732
Conc: 86.30 ng/ml



#35 AR-1260-5


R.T.: 8.373 min
Delta R.T.: -0.005 min
Response: 2335910
Conc: 18.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#35 AR-1260-5

R.T.: 7.321 min
Delta R.T.: -0.029 min
Response: 14420772
Conc: 98.33 ng/ml




#36 AR-1262-1

R.T.: 7.813 min
Delta R.T.: -0.015 min
Response: 11857957
Conc: 136.37 ng/ml



#36 AR-1262-1


R.T.: 6.881 min
Delta R.T.: -0.018 min
Response: 7005137
Conc: 186.40 ng/ml



#37 AR-1262-2


R.T.: 8.373 min
Delta R.T.: -0.004 min
Response: 2335910
Conc: 15.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#37 AR-1262-2

R.T.: 7.101 min
Delta R.T.: -0.006 min
Response: 5432732
Conc: 64.14 ng/ml




#38 AR-1262-3

R.T.: 8.678 min
Delta R.T.: -0.019 min
Response: 1724863
Conc: 16.75 ng/ml



#38 AR-1262-3


R.T.: 7.649 min
Delta R.T.: 0.013 min
Response: 625351
Conc: 8.92 ng/ml



#39 AR-1262-4


R.T.: 8.789 min
Delta R.T.: 0.002 min
Response: 12322615
Conc: 267.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#39 AR-1262-4

R.T.: 7.723 min
Delta R.T.: 0.022 min
Response: 11553474
Conc: 92.47 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.459 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.215 min
Delta R.T.: 0.013 min
Response: 7639322
Conc: 134.83 ng/ml



#41 AR-1268-1


R.T.: 8.678 min
Delta R.T.: -0.015 min
Response: 1724863
Conc: 9.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#41 AR-1268-1

R.T.: 7.649 min
Delta R.T.: 0.013 min
Response: 625351
Conc: 3.15 ng/ml




#42 AR-1268-2

R.T.: 8.789 min
Delta R.T.: -0.001 min
Response: 12322615
Conc: 74.33 ng/ml



#42 AR-1268-2


R.T.: 7.723 min
Delta R.T.: 0.022 min
Response: 11553474
Conc: 64.65 ng/ml



#43 AR-1268-3


R.T.: 9.025 min
Delta R.T.: 0.000 min
Response: 358108
Conc: 2.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
CLI-0120




#43 AR-1268-3

R.T.: 7.906 min
Delta R.T.: -0.004 min
Response: 497544
Conc: 3.25 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.460 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 8.215 min
Delta R.T.: 0.014 min
Response: 7639322
Conc: 121.71 ng/ml