

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
 Data File : PP051957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 08:52
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 01:52:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.675	82873736	64304074	46.368	50.580
2) SA Decachlor...	10.243	8.770	62964414	49821823	51.938	42.182
Target Compounds						
3) L1 AR-1016-1	5.598	4.778	21325479	22010932	470.544	501.939
4) L1 AR-1016-2	5.620	4.798	30051684	30968904	468.222	501.970
5) L1 AR-1016-3	5.683	4.977	19492613	16760004	487.628	510.078
6) L1 AR-1016-4	5.782	5.019	16793813	12256527	479.036	473.911
7) L1 AR-1016-5	6.079	5.236	19824951	19075956	480.045	553.920
31) L7 AR-1260-1	7.211	6.285	40038459	33528365	563.281	486.767
32) L7 AR-1260-2	7.468	6.474	42435863	39231275	519.562	466.188
33) L7 AR-1260-3	7.829	6.630	29727157	35382159	519.757	454.526
34) L7 AR-1260-4	8.056	7.108	33454894	28576605	510.805	453.942
35) L7 AR-1260-5	8.380	7.350	62250292	62469338	489.562	425.938



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
Data File : PP051957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 08:52
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 01:52:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: 0.001 min
Response: 82873736
Conc: 46.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: 0.000 min
Response: 64304074
Conc: 50.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: 0.003 min
Response: 62964414
Conc: 51.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 49821823
Conc: 42.18 ng/ml



#3 AR-1016-1


R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 21325479
Conc: 470.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 22010932
Conc: 501.94 ng/ml




#4 AR-1016-2

R.T.: 5.620 min
Delta R.T.: 0.000 min
Response: 30051684
Conc: 468.22 ng/ml



#4 AR-1016-2


R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 30968904
Conc: 501.97 ng/ml



#5 AR-1016-3


R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 19492613
Conc: 487.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.977 min
Delta R.T.: 0.000 min
Response: 16760004
Conc: 510.08 ng/ml




#6 AR-1016-4

R.T.: 5.782 min
Delta R.T.: 0.001 min
Response: 16793813
Conc: 479.04 ng/ml



#6 AR-1016-4


R.T.: 5.019 min
Delta R.T.: 0.000 min
Response: 12256527
Conc: 473.91 ng/ml



#7 AR-1016-5


R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 19824951
Conc: 480.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 19075956
Conc: 553.92 ng/ml




#31 AR-1260-1

R.T.: 7.211 min
Delta R.T.: 0.000 min
Response: 40038459
Conc: 563.28 ng/ml



#31 AR-1260-1


R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 33528365
Conc: 486.77 ng/ml



#32 AR-1260-2


R.T.: 7.468 min
Delta R.T.: 0.000 min
Response: 42435863
Conc: 519.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 39231275
Conc: 466.19 ng/ml




#33 AR-1260-3

R.T.: 7.829 min
Delta R.T.: 0.000 min
Response: 29727157
Conc: 519.76 ng/ml



#33 AR-1260-3


R.T.: 6.630 min
Delta R.T.: 0.000 min
Response: 35382159
Conc: 454.53 ng/ml



#34 AR-1260-4


R.T.: 8.056 min
Delta R.T.: 0.002 min
Response: 33454894
Conc: 510.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.108 min
Delta R.T.: 0.000 min
Response: 28576605
Conc: 453.94 ng/ml



#35 AR-1260-5

R.T.: 8.380 min
Delta R.T.: 0.002 min
Response: 62250292
Conc: 489.56 ng/ml



#35 AR-1260-5

R.T.: 7.350 min
Delta R.T.: 0.000 min
Response: 62469338
Conc: 425.94 ng/ml