

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
 Data File : PP051960.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 09:41
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 01:54:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.675	87389231	69799968	48.895	54.903
2) SA Decachlor...	10.244	8.771	67631378	54174111	55.788	45.866
Target Compounds						
26) L6 AR-1254-1	6.459	5.594	33620185	37622472	492.715	505.719
27) L6 AR-1254-2	6.678	5.744	48230230	32040450	471.296	496.861
28) L6 AR-1254-3	7.046	6.153	48749930	50876359	476.144	483.231
29) L6 AR-1254-4	7.332	6.384	35636893	36877030	496.777	540.993
30) L6 AR-1254-5	7.753	6.807	37526442	41794217	498.809	451.755



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
Data File : PP051960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 09:41
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 01:54:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: 0.001 min
Response: 87389231
Conc: 48.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: 0.001 min
Response: 69799968
Conc: 54.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: 0.003 min
Response: 67631378
Conc: 55.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.771 min
Delta R.T.: 0.000 min
Response: 54174111
Conc: 45.87 ng/ml



#26 AR-1254-1


R.T.: 6.459 min
Delta R.T.: 0.002 min
Response: 33620185
Conc: 492.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 37622472
Conc: 505.72 ng/ml




#27 AR-1254-2

R.T.: 6.678 min
Delta R.T.: 0.002 min
Response: 48230230
Conc: 471.30 ng/ml



#27 AR-1254-2


R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 32040450
Conc: 496.86 ng/ml



#28 AR-1254-3


R.T.: 7.046 min
Delta R.T.: 0.002 min
Response: 48749930
Conc: 476.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 50876359
Conc: 483.23 ng/ml




#29 AR-1254-4

R.T.: 7.332 min
Delta R.T.: 0.002 min
Response: 35636893
Conc: 496.78 ng/ml



#29 AR-1254-4


R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 36877030
Conc: 540.99 ng/ml



#30 AR-1254-5

R.T.: 7.753 min
Delta R.T.: 0.002 min
Response: 37526442
Conc: 498.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 41794217
Conc: 451.75 ng/ml