

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
 Data File : PP051969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 12:11
 Operator : YP\AJ
 Sample : N4957-12
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-28

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 01:59:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.676	42778123	28772263	23.935	22.631
2) SA Decachlor...	10.243	8.771	35947481	26069184	29.652	22.071 #
Target Compounds						
21) L5 AR-1248-1	5.598	4.779	542.4E6	542.3E6	17344.376	17957.765
22) L5 AR-1248-2	5.874	5.020	1319.8E6	1086.4E6	25003.270	26120.041
23) L5 AR-1248-3	6.080	5.063	2286.3E6	1362.3E6	36960.079	29115.166
24) L5 AR-1248-4	6.485	5.238	1745.3E6	1832.8E6	25317.285	35754.119 #
25) L5 AR-1248-5	6.524	5.637	1630.2E6	1153.2E6	23886.038	21701.577



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100522\
Data File : PP051969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 12:11
Operator : YP\AJ
Sample : N4957-12
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB-28

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 01:59:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: 0.001 min
Response: 42778123
Conc: 23.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-28




#1 Tetrachloro-m-xylene

R.T.: 3.676 min
Delta R.T.: 0.002 min
Response: 28772263
Conc: 22.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: 0.003 min
Response: 35947481
Conc: 29.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.771 min
Delta R.T.: 0.000 min
Response: 26069184
Conc: 22.07 ng/ml



#21 AR-1248-1


R.T.: 5.598 min
Delta R.T.: 0.002 min
Response: 542411758
Conc: 17344.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-28




#21 AR-1248-1

R.T.: 4.779 min
Delta R.T.: 0.001 min
Response: 542304177
Conc: 17957.76 ng/ml




#22 AR-1248-2

R.T.: 5.874 min
Delta R.T.: 0.002 min
Response: 1319781471
Conc: 25003.27 ng/ml



#22 AR-1248-2


R.T.: 5.020 min
Delta R.T.: 0.001 min
Response: 1086395355
Conc: 26120.04 ng/ml



#23 AR-1248-3


R.T.: 6.080 min
Delta R.T.: 0.003 min
Response: 2286279419
Conc: 36960.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-28




#23 AR-1248-3

R.T.: 5.063 min
Delta R.T.: 0.002 min
Response: 1362272554
Conc: 29115.17 ng/ml




#24 AR-1248-4

R.T.: 6.485 min
Delta R.T.: 0.002 min
Response: 1745308068
Conc: 25317.28 ng/ml



#24 AR-1248-4


R.T.: 5.238 min
Delta R.T.: 0.002 min
Response: 1832819691
Conc: 35754.12 ng/ml



#25 AR-1248-5

R.T.: 6.524 min
Delta R.T.: 0.001 min
Response: 1630184866
Conc: 23886.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-28



#25 AR-1248-5

R.T.: 5.637 min
Delta R.T.: 0.001 min
Response: 1153191282
Conc: 21701.58 ng/ml