

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100523\
 Data File : PP060668.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2023 04:57
 Operator : YP\AJ
 Sample : 04763-01
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PAT-1-100423

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/06/2023
 Supervised By :Ankita Jodhani 10/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 05:19:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	44447131	31022972	19.847	20.842
2) SA Decachlor...	10.458	8.747	33915517	25090771	19.210	15.382

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100523\
Data File : PP060668.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2023 04:57
Operator : YP\AJ
Sample : 04763-01
Misc :
ALS Vial : 49 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PAT-1-100423


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/06/2023
Supervised By :Ankita Jodhani 10/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 05:19:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 44447131
Conc: 19.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAT-1-100423


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/06/2023
Supervised By :Ankita Jodhani 10/06/2023




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 31022972
Conc: 20.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.458 min
Delta R.T.: 0.002 min
Response: 33915517
Conc: 19.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.747 min
Delta R.T.: -0.003 min
Response: 25090771
Conc: 15.38 ng/ml