

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100621\
 Data File : PP039857.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2021 18:31
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP100621AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 18:29:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 17:56:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.899	3.987	1554243	860277	51.051	51.463
2) SA Decachlor...	10.924	9.319	1144768	1116322	50.921	49.763
Target Compounds						
26) L6 AR-1254-1	7.135	6.130	579470	491768	562.629	498.999
27) L6 AR-1254-2	7.365	6.290	814143	451139	506.492	495.420
28) L6 AR-1254-3	7.748	6.713	849143	708507	500.816	504.436
29) L6 AR-1254-4	8.045	6.952	602781	445070	496.099	504.846
30) L6 AR-1254-5	8.473	7.384	634873	673555	499.111	502.897



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100621\
Data File : PP039857.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2021 18:31
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP100621AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 18:29:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 17:56:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.899 min
Delta R.T.: 0.002 min
Response: 1554243
Conc: 51.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100621AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.987 min
Delta R.T.: 0.003 min
Response: 860277
Conc: 51.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.924 min
Delta R.T.: 0.004 min
Response: 1144768
Conc: 50.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.319 min
Delta R.T.: 0.002 min
Response: 1116322
Conc: 49.76 ng/ml



#26 AR-1254-1


R.T.: 7.135 min
Delta R.T.: 0.002 min
Response: 579470
Conc: 562.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100621AR1254




#26 AR-1254-1

R.T.: 6.130 min
Delta R.T.: 0.002 min
Response: 491768
Conc: 499.00 ng/ml




#27 AR-1254-2

R.T.: 7.365 min
Delta R.T.: 0.001 min
Response: 814143
Conc: 506.49 ng/ml



#27 AR-1254-2


R.T.: 6.290 min
Delta R.T.: 0.001 min
Response: 451139
Conc: 495.42 ng/ml



#28 AR-1254-3


R.T.: 7.748 min
Delta R.T.: 0.002 min
Response: 849143
Conc: 500.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100621AR1254




#28 AR-1254-3

R.T.: 6.713 min
Delta R.T.: 0.001 min
Response: 708507
Conc: 504.44 ng/ml




#29 AR-1254-4

R.T.: 8.045 min
Delta R.T.: 0.002 min
Response: 602781
Conc: 496.10 ng/ml



#29 AR-1254-4


R.T.: 6.952 min
Delta R.T.: 0.000 min
Response: 445070
Conc: 504.85 ng/ml



#30 AR-1254-5

R.T.: 8.473 min
Delta R.T.: 0.003 min
Response: 634873
Conc: 499.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100621AR1254



#30 AR-1254-5

R.T.: 7.384 min
Delta R.T.: 0.001 min
Response: 673555
Conc: 502.90 ng/ml