

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100622\
 Data File : PP052062.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 16:17
 Operator : YP\AJ
 Sample : N5008-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 2007

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/07/2022
 Supervised By :Ankita Jodhani 10/07/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 17:10:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.687	24328460	16093765	13.612	12.659m
2) SA Decachlor...	10.244	8.772	30347159	24504556	25.033	20.747

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100622\
Data File : PP052062.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 16:17
Operator : YP\AJ
Sample : N5008-04
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
2007


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/07/2022
Supervised By :Ankita Jodhani 10/07/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 17:10:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.432 min
Delta R.T.: 0.009 min
Response: 24328460
Conc: 13.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
2007


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/07/2022
Supervised By :Ankita Jodhani 10/07/2022




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: 0.012 min
Response: 16093765
Conc: 12.66 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: 0.004 min
Response: 30347159
Conc: 25.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.772 min
Delta R.T.: 0.002 min
Response: 24504556
Conc: 20.75 ng/ml