

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100622\
 Data File : PP052070.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2022 19:37
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 00:22:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.425	3.676	89969712	71065532	50.339	55.898
2) SA Decachlor...	10.244	8.770	74475025	60104299	61.433	50.887
Target Compounds						
26) L6 AR-1254-1	6.459	5.595	36123424	37238990	529.400	500.564
27) L6 AR-1254-2	6.679	5.744	54292841	32203630	530.538	499.392
28) L6 AR-1254-3	7.047	6.154	53448995	52126133	522.040	495.101
29) L6 AR-1254-4	7.333	6.384	37094337	31904642	517.094	468.047
30) L6 AR-1254-5	7.754	6.808	40503295	44812140	538.378	484.375



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100622\
Data File : PP052070.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2022 19:37
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 00:22:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.002 min
Response: 89969712
Conc: 50.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.676 min
Delta R.T.: 0.002 min
Response: 71065532
Conc: 55.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.244 min
Delta R.T.: 0.004 min
Response: 74475025
Conc: 61.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 60104299
Conc: 50.89 ng/ml



#26 AR-1254-1


R.T.: 6.459 min
Delta R.T.: 0.002 min
Response: 36123424
Conc: 529.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.595 min
Delta R.T.: 0.002 min
Response: 37238990
Conc: 500.56 ng/ml




#27 AR-1254-2

R.T.: 6.679 min
Delta R.T.: 0.003 min
Response: 54292841
Conc: 530.54 ng/ml



#27 AR-1254-2


R.T.: 5.744 min
Delta R.T.: 0.002 min
Response: 32203630
Conc: 499.39 ng/ml



#28 AR-1254-3


R.T.: 7.047 min
Delta R.T.: 0.003 min
Response: 53448995
Conc: 522.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.154 min
Delta R.T.: 0.002 min
Response: 52126133
Conc: 495.10 ng/ml




#29 AR-1254-4

R.T.: 7.333 min
Delta R.T.: 0.003 min
Response: 37094337
Conc: 517.09 ng/ml



#29 AR-1254-4


R.T.: 6.384 min
Delta R.T.: 0.001 min
Response: 31904642
Conc: 468.05 ng/ml



#30 AR-1254-5

R.T.: 7.754 min
Delta R.T.: 0.003 min
Response: 40503295
Conc: 538.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.808 min
Delta R.T.: 0.001 min
Response: 44812140
Conc: 484.38 ng/ml