

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100623\
 Data File : PP060736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2023 03:25
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 04:35:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.532	3.684	115.5E6	81498306	51.582	54.752m
2) SA Decachlor...	10.461	8.747	88051458	76662568	49.872	46.998
Target Compounds						
16) L4 AR-1242-1	5.719	4.781	32828898	25039388	513.697	494.584
17) L4 AR-1242-2	5.742	4.801	46082557	34513642	508.939	496.997
18) L4 AR-1242-3	5.806	4.978	29106396	19340076	516.982	502.200
19) L4 AR-1242-4	5.906	5.063	24338191	19126562	521.094	485.656
20) L4 AR-1242-5	6.653	5.590	27627313	23917308	549.295	488.935



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100623\
Data File : PP060736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2023 03:25
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 04:35:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.532 min
Delta R.T.: 0.000 min
Response: 115518331
Conc: 51.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.001 min
Response: 81498306
Conc: 54.75 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.461 min
Delta R.T.: 0.006 min
Response: 88051458
Conc: 49.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: -0.003 min
Response: 76662568
Conc: 47.00 ng/ml



#16 AR-1242-1


R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 32828898
Conc: 513.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.781 min
Delta R.T.: -0.003 min
Response: 25039388
Conc: 494.58 ng/ml




#17 AR-1242-2

R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 46082557
Conc: 508.94 ng/ml



#17 AR-1242-2


R.T.: 4.801 min
Delta R.T.: -0.003 min
Response: 34513642
Conc: 497.00 ng/ml



#18 AR-1242-3


R.T.: 5.806 min
Delta R.T.: 0.000 min
Response: 29106396
Conc: 516.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.978 min
Delta R.T.: -0.003 min
Response: 19340076
Conc: 502.20 ng/ml




#19 AR-1242-4

R.T.: 5.906 min
Delta R.T.: 0.000 min
Response: 24338191
Conc: 521.09 ng/ml



#19 AR-1242-4


R.T.: 5.063 min
Delta R.T.: -0.003 min
Response: 19126562
Conc: 485.66 ng/ml



#20 AR-1242-5

R.T.: 6.653 min
Delta R.T.: 0.000 min
Response: 27627313
Conc: 549.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.590 min
Delta R.T.: -0.003 min
Response: 23917308
Conc: 488.93 ng/ml