

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100721\
 Data File : PP039875.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Oct 2021 23:47
 Operator : AJ\MA
 Sample : M4119-03MSD
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 20-TW-2-20211006MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 05:18:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 08 04:46:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.898	3.985	740563	367918	24.324	22.009
2)	SA Decachlor...	10.920	9.317	341841	342241	15.206	15.256
Target Compounds							
3)	L1 AR-1016-1	6.222	5.251	575925	364071	558.948	538.190
4)	L1 AR-1016-2	6.246	5.271	831608	496114	557.069	536.240
5)	L1 AR-1016-3	6.312	5.464	526568	279015	565.219	535.365
6)	L1 AR-1016-4	6.419	5.516	433621	221856	565.805	512.430
7)	L1 AR-1016-5	6.735	5.747	412926	285168	546.032	543.770
31)	L7 AR-1260-1	7.913	6.848	651315	516162	494.099	491.581
32)	L7 AR-1260-2	8.179	7.045	760055	603095	473.810	477.269
33)	L7 AR-1260-3	8.545	7.202	414687	594130	424.653	448.086
34)	L7 AR-1260-4	8.776	7.687	503383	417919	452.040	450.441
35)	L7 AR-1260-5	9.101	7.935	931461	950029	457.379	455.200



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100721\
Data File : PP039875.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Oct 2021 23:47
Operator : AJ\MA
Sample : M4119-03MSD
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 05:18:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 08 04:46:38 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.898 min
Delta R.T.: 0.001 min
Response: 740563
Conc: 24.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: 0.000 min
Response: 367918
Conc: 22.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.920 min
Delta R.T.: 0.003 min
Response: 341841
Conc: 15.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.317 min
Delta R.T.: 0.000 min
Response: 342241
Conc: 15.26 ng/ml



#3 AR-1016-1


R.T.: 6.222 min
Delta R.T.: 0.001 min
Response: 575925
Conc: 558.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#3 AR-1016-1

R.T.: 5.251 min
Delta R.T.: 0.000 min
Response: 364071
Conc: 538.19 ng/ml




#4 AR-1016-2

R.T.: 6.246 min
Delta R.T.: 0.001 min
Response: 831608
Conc: 557.07 ng/ml



#4 AR-1016-2


R.T.: 5.271 min
Delta R.T.: 0.000 min
Response: 496114
Conc: 536.24 ng/ml



#5 AR-1016-3


R.T.: 6.312 min
Delta R.T.: 0.000 min
Response: 526568
Conc: 565.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#5 AR-1016-3

R.T.: 5.464 min
Delta R.T.: 0.000 min
Response: 279015
Conc: 535.37 ng/ml




#6 AR-1016-4

R.T.: 6.419 min
Delta R.T.: 0.000 min
Response: 433621
Conc: 565.81 ng/ml



#6 AR-1016-4


R.T.: 5.516 min
Delta R.T.: 0.000 min
Response: 221856
Conc: 512.43 ng/ml



#7 AR-1016-5


R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 412926
Conc: 546.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#7 AR-1016-5

R.T.: 5.747 min
Delta R.T.: 0.000 min
Response: 285168
Conc: 543.77 ng/ml




#31 AR-1260-1

R.T.: 7.913 min
Delta R.T.: 0.001 min
Response: 651315
Conc: 494.10 ng/ml



#31 AR-1260-1


R.T.: 6.848 min
Delta R.T.: 0.000 min
Response: 516162
Conc: 491.58 ng/ml



#32 AR-1260-2


R.T.: 8.179 min
Delta R.T.: 0.000 min
Response: 760055
Conc: 473.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#32 AR-1260-2

R.T.: 7.045 min
Delta R.T.: 0.000 min
Response: 603095
Conc: 477.27 ng/ml




#33 AR-1260-3

R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 414687
Conc: 424.65 ng/ml



#33 AR-1260-3


R.T.: 7.202 min
Delta R.T.: 0.000 min
Response: 594130
Conc: 448.09 ng/ml



#34 AR-1260-4


R.T.: 8.776 min
Delta R.T.: 0.000 min
Response: 503383
Conc: 452.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
20-TW-2-20211006MSD




#34 AR-1260-4

R.T.: 7.687 min
Delta R.T.: 0.000 min
Response: 417919
Conc: 450.44 ng/ml



#35 AR-1260-5

R.T.: 9.101 min
Delta R.T.: 0.001 min
Response: 931461
Conc: 457.38 ng/ml



#35 AR-1260-5

R.T.: 7.935 min
Delta R.T.: 0.001 min
Response: 950029
Conc: 455.20 ng/ml