

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100722\
 Data File : PP052122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2022 15:25
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 17:39:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 04 12:08:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.677	77387588	61036041	43.299	48.009
2) SA Decachlor...	10.249	8.775	60943927	48677815	50.271	41.213
Target Compounds						
3) L1 AR-1016-1	5.602	4.781	19554857	20554464	431.475	468.726
4) L1 AR-1016-2	5.624	4.800	27594337	28926297	429.935	468.862
5) L1 AR-1016-3	5.687	4.979	17864530	15627780	446.899	475.620
6) L1 AR-1016-4	5.785	5.021	15698160	12562270	447.783	485.733
7) L1 AR-1016-5	6.083	5.239	18844433	16354150	456.302	474.886
31) L7 AR-1260-1	7.215	6.287	33904899	30314063	476.991	440.101
32) L7 AR-1260-2	7.472	6.477	37668140	36041851	461.189	428.288
33) L7 AR-1260-3	7.833	6.633	28411068	33410025	496.746	429.192
34) L7 AR-1260-4	8.060	7.111	32202930	26865570	491.690	426.762
35) L7 AR-1260-5	8.385	7.353	60777126	59088971	477.977	402.889



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100722\
Data File : PP052122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2022 15:25
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 17:39:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100322.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 04 12:08:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.427 min
Delta R.T.: 0.004 min
Response: 77387588
Conc: 43.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.677 min
Delta R.T.: 0.003 min
Response: 61036041
Conc: 48.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.249 min
Delta R.T.: 0.009 min
Response: 60943927
Conc: 50.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.775 min
Delta R.T.: 0.005 min
Response: 48677815
Conc: 41.21 ng/ml



#3 AR-1016-1


R.T.: 5.602 min
Delta R.T.: 0.005 min
Response: 19554857
Conc: 431.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.781 min
Delta R.T.: 0.003 min
Response: 20554464
Conc: 468.73 ng/ml




#4 AR-1016-2

R.T.: 5.624 min
Delta R.T.: 0.004 min
Response: 27594337
Conc: 429.94 ng/ml



#4 AR-1016-2


R.T.: 4.800 min
Delta R.T.: 0.003 min
Response: 28926297
Conc: 468.86 ng/ml



#5 AR-1016-3


R.T.: 5.687 min
Delta R.T.: 0.004 min
Response: 17864530
Conc: 446.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.979 min
Delta R.T.: 0.003 min
Response: 15627780
Conc: 475.62 ng/ml




#6 AR-1016-4

R.T.: 5.785 min
Delta R.T.: 0.004 min
Response: 15698160
Conc: 447.78 ng/ml



#6 AR-1016-4


R.T.: 5.021 min
Delta R.T.: 0.002 min
Response: 12562270
Conc: 485.73 ng/ml



#7 AR-1016-5


R.T.: 6.083 min
Delta R.T.: 0.004 min
Response: 18844433
Conc: 456.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.239 min
Delta R.T.: 0.003 min
Response: 16354150
Conc: 474.89 ng/ml




#31 AR-1260-1

R.T.: 7.215 min
Delta R.T.: 0.005 min
Response: 33904899
Conc: 476.99 ng/ml



#31 AR-1260-1


R.T.: 6.287 min
Delta R.T.: 0.002 min
Response: 30314063
Conc: 440.10 ng/ml



#32 AR-1260-2


R.T.: 7.472 min
Delta R.T.: 0.005 min
Response: 37668140
Conc: 461.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.477 min
Delta R.T.: 0.003 min
Response: 36041851
Conc: 428.29 ng/ml




#33 AR-1260-3

R.T.: 7.833 min
Delta R.T.: 0.005 min
Response: 28411068
Conc: 496.75 ng/ml



#33 AR-1260-3


R.T.: 6.633 min
Delta R.T.: 0.003 min
Response: 33410025
Conc: 429.19 ng/ml



#34 AR-1260-4


R.T.: 8.060 min
Delta R.T.: 0.005 min
Response: 32202930
Conc: 491.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.111 min
Delta R.T.: 0.003 min
Response: 26865570
Conc: 426.76 ng/ml



#35 AR-1260-5

R.T.: 8.385 min
Delta R.T.: 0.006 min
Response: 60777126
Conc: 477.98 ng/ml



#35 AR-1260-5

R.T.: 7.353 min
Delta R.T.: 0.003 min
Response: 59088971
Conc: 402.89 ng/ml