

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
 Data File : PP067548.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2024 14:44
 Operator : YP\AJ
 Sample : P4324-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LAW-24-00169-00170

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:06:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.756	4.055	13452909	13505781	10.009	9.039
2)	SA Decachlor...	10.669	9.221	17803986	16001400	12.264	11.612
Target Compounds							
3)	L1 AR-1016-1	5.892	5.168	482871	379809	10.127	7.364 #
4)	L1 AR-1016-2	5.979f	5.197	547545	328357	7.908	4.618 #
5)	L1 AR-1016-3	6.023	5.357	384038	338262	8.597	8.500
6)	L1 AR-1016-4	6.064f	5.405	498066	432268	13.860	12.519
7)	L1 AR-1016-5	0.000	5.621	0	145351	N.D.	3.387 #
8)	L2 AR-1221-1	4.957	4.229f	1754062	173519	104.778	9.076 #
9)	L2 AR-1221-2	0.000	4.363	0	269001	N.D.	18.335 #
10)	L2 AR-1221-3	5.145	4.434	448583	283064	11.591	6.528 #
11)	L3 AR-1232-1	5.145	4.434	448583	283064	13.991	7.869 #
12)	L3 AR-1232-2	5.610f	5.197	1336478	328357	75.085	9.912 #
13)	L3 AR-1232-3	5.979f	5.357	547545	338262	17.320	18.828
14)	L3 AR-1232-4	6.064f	5.438	498066	63429	30.865	3.627 #
15)	L3 AR-1232-5	6.209	5.621	335194	145351	24.777	7.741 #
16)	L4 AR-1242-1	5.892	5.168	482871	379809	13.578	9.905 #
17)	L4 AR-1242-2	5.979f	5.197	547545	328357	10.465	6.270 #
18)	L4 AR-1242-3	6.023	5.357	384038	338262	11.268	11.381
19)	L4 AR-1242-4	6.064f	5.438	498066	63429	18.160	1.971 #
20)	L4 AR-1242-5	6.851	0.000	374121	0	12.571	N.D. #
21)	L5 AR-1248-1	5.892	5.168	482871	379809	17.102	12.402 #
22)	L5 AR-1248-2	6.209	5.405	335194	432268	7.550	9.528 #
23)	L5 AR-1248-3	0.000	5.438	0	63429	N.D.	1.340 #
24)	L5 AR-1248-4	6.769	5.621	246403	145351	4.541	2.676 #
25)	L5 AR-1248-5	6.851	6.074f	374121	176815	6.784	3.628 #
26)	L6 AR-1254-1	6.769	0.000	246403	0	3.955	N.D. #
27)	L6 AR-1254-2	7.002	6.128	1143785	148365	12.606	2.023 #
28)	L6 AR-1254-3	7.356	6.539	597601	723243	6.371	6.343
29)	L6 AR-1254-4	7.643	6.760	787369	62540	11.815	1.002 #
30)	L6 AR-1254-5	8.070	7.181	1651344	1434348	20.090	13.971 #
31)	L7 AR-1260-1	7.525	6.664	407129	356757	5.168	4.361
32)	L7 AR-1260-2	7.773	6.851	1433987	543073	15.733	5.781 #
33)	L7 AR-1260-3	8.142	7.008	284434	240053	4.503	2.650 #
34)	L7 AR-1260-4	8.378	7.484	202028	155645	2.676	2.334
35)	L7 AR-1260-5	8.710	7.724	2121017	1134962	16.382	7.796 #
36)	L8 AR-1262-1	8.378	7.221	202028	177731	2.119	1.690
37)	L8 AR-1262-2	8.710	7.484	2121017	155645	13.073	1.644 #
38)	L8 AR-1262-3	9.045	8.008	421555	147519	3.588	2.011 #
39)	L8 AR-1262-4	9.135	8.074	461052	472163	4.985	3.582 #
40)	L8 AR-1262-5	0.000	8.637	0	134585	N.D.	2.215 #
41)	L9 AR-1268-1	9.045	8.008	421555	147519	2.190	0.732 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
 Data File : PP067548.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2024 14:44
 Operator : YP\AJ
 Sample : P4324-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LAW-24-00169-00170

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:06:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.135	8.074	461052	472163	2.703	2.640
43) L9	AR-1268-3	0.000	8.326	0	305451	N.D.	1.923 #
44) L9	AR-1268-4	9.769f	8.637	416333	134585	6.842	2.085 #
45) L9	AR-1268-5	10.326f	8.925	1247772	742006	2.888	1.680 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
Data File : PP067548.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2024 14:44
Operator : YP\AJ
Sample : P4324-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 02:06:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.756 min
Delta R.T.: 0.003 min
Response: 13452909
Conc: 10.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#1 Tetrachloro-m-xylene

R.T.: 4.055 min
Delta R.T.: 0.002 min
Response: 13505781
Conc: 9.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.669 min
Delta R.T.: 0.007 min
Response: 17803986
Conc: 12.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.221 min
Delta R.T.: -0.005 min
Response: 16001400
Conc: 11.61 ng/ml



#3 AR-1016-1


R.T.: 5.892 min
Delta R.T.: -0.024 min
Response: 482871
Conc: 10.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#3 AR-1016-1

R.T.: 5.168 min
Delta R.T.: 0.004 min
Response: 379809
Conc: 7.36 ng/ml




#4 AR-1016-2

R.T.: 5.979 min
Delta R.T.: 0.039 min
Response: 547545
Conc: 7.91 ng/ml



#4 AR-1016-2


R.T.: 5.197 min
Delta R.T.: 0.013 min
Response: 328357
Conc: 4.62 ng/ml



#5 AR-1016-3


R.T.: 6.023 min
Delta R.T.: 0.019 min
Response: 384038
Conc: 8.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#5 AR-1016-3

R.T.: 5.357 min
Delta R.T.: -0.008 min
Response: 338262
Conc: 8.50 ng/ml




#6 AR-1016-4

R.T.: 6.064 min
Delta R.T.: -0.037 min
Response: 498066
Conc: 13.86 ng/ml



#6 AR-1016-4


R.T.: 5.405 min
Delta R.T.: 0.000 min
Response: 432268
Conc: 12.52 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.396 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#7 AR-1016-5

R.T.: 5.621 min
Delta R.T.: -0.003 min
Response: 145351
Conc: 3.39 ng/ml




#8 AR-1221-1

R.T.: 4.957 min
Delta R.T.: 0.002 min
Response: 1754062
Conc: 104.78 ng/ml



#8 AR-1221-1


R.T.: 4.229 min
Delta R.T.: -0.040 min
Response: 173519
Conc: 9.08 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 5.041 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#9 AR-1221-2

R.T.: 4.363 min
Delta R.T.: 0.006 min
Response: 269001
Conc: 18.34 ng/ml




#10 AR-1221-3

R.T.: 5.145 min
Delta R.T.: 0.027 min
Response: 448583
Conc: 11.59 ng/ml



#10 AR-1221-3


R.T.: 4.434 min
Delta R.T.: -0.002 min
Response: 283064
Conc: 6.53 ng/ml



#11 AR-1232-1


R.T.: 5.145 min
Delta R.T.: 0.028 min
Response: 448583
Conc: 13.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#11 AR-1232-1

R.T.: 4.434 min
Delta R.T.: -0.001 min
Response: 283064
Conc: 7.87 ng/ml




#12 AR-1232-2

R.T.: 5.610 min
Delta R.T.: -0.040 min
Response: 1336478
Conc: 75.09 ng/ml



#12 AR-1232-2


R.T.: 5.197 min
Delta R.T.: 0.014 min
Response: 328357
Conc: 9.91 ng/ml



#13 AR-1232-3


R.T.: 5.979 min
Delta R.T.: 0.040 min
Response: 547545
Conc: 17.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#13 AR-1232-3

R.T.: 5.357 min
Delta R.T.: -0.007 min
Response: 338262
Conc: 18.83 ng/ml




#14 AR-1232-4

R.T.: 6.064 min
Delta R.T.: -0.036 min
Response: 498066
Conc: 30.87 ng/ml



#14 AR-1232-4


R.T.: 5.438 min
Delta R.T.: -0.009 min
Response: 63429
Conc: 3.63 ng/ml



#15 AR-1232-5


R.T.: 6.209 min
Delta R.T.: 0.019 min
Response: 335194
Conc: 24.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#15 AR-1232-5

R.T.: 5.621 min
Delta R.T.: -0.001 min
Response: 145351
Conc: 7.74 ng/ml




#16 AR-1242-1

R.T.: 5.892 min
Delta R.T.: -0.026 min
Response: 482871
Conc: 13.58 ng/ml



#16 AR-1242-1


R.T.: 5.168 min
Delta R.T.: 0.004 min
Response: 379809
Conc: 9.90 ng/ml



#17 AR-1242-2


R.T.: 5.979 min
Delta R.T.: 0.038 min
Response: 547545
Conc: 10.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#17 AR-1242-2

R.T.: 5.197 min
Delta R.T.: 0.013 min
Response: 328357
Conc: 6.27 ng/ml




#18 AR-1242-3

R.T.: 6.023 min
Delta R.T.: 0.019 min
Response: 384038
Conc: 11.27 ng/ml



#18 AR-1242-3


R.T.: 5.357 min
Delta R.T.: -0.008 min
Response: 338262
Conc: 11.38 ng/ml



#19 AR-1242-4


R.T.: 6.064 min
Delta R.T.: -0.038 min
Response: 498066
Conc: 18.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#19 AR-1242-4

R.T.: 5.438 min
Delta R.T.: -0.010 min
Response: 63429
Conc: 1.97 ng/ml




#20 AR-1242-5

R.T.: 6.851 min
Delta R.T.: 0.015 min
Response: 374121
Conc: 12.57 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 5.978 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.892 min
Delta R.T.: -0.024 min
Response: 482871
Conc: 17.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#21 AR-1248-1

R.T.: 5.168 min
Delta R.T.: 0.005 min
Response: 379809
Conc: 12.40 ng/ml




#22 AR-1248-2

R.T.: 6.209 min
Delta R.T.: 0.019 min
Response: 335194
Conc: 7.55 ng/ml



#22 AR-1248-2


R.T.: 5.405 min
Delta R.T.: 0.001 min
Response: 432268
Conc: 9.53 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.395 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#23 AR-1248-3

R.T.: 5.438 min
Delta R.T.: -0.009 min
Response: 63429
Conc: 1.34 ng/ml




#24 AR-1248-4

R.T.: 6.769 min
Delta R.T.: -0.025 min
Response: 246403
Conc: 4.54 ng/ml



#24 AR-1248-4


R.T.: 5.621 min
Delta R.T.: -0.002 min
Response: 145351
Conc: 2.68 ng/ml



#25 AR-1248-5


R.T.: 6.851 min
Delta R.T.: 0.016 min
Response: 374121
Conc: 6.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#25 AR-1248-5

R.T.: 6.074 min
Delta R.T.: 0.053 min
Response: 176815
Conc: 3.63 ng/ml




#26 AR-1254-1

R.T.: 6.769 min
Delta R.T.: -0.001 min
Response: 246403
Conc: 3.95 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 5.980 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 7.002 min
Delta R.T.: 0.014 min
Response: 1143785
Conc: 12.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#27 AR-1254-2

R.T.: 6.128 min
Delta R.T.: 0.000 min
Response: 148365
Conc: 2.02 ng/ml




#28 AR-1254-3

R.T.: 7.356 min
Delta R.T.: 0.004 min
Response: 597601
Conc: 6.37 ng/ml



#28 AR-1254-3


R.T.: 6.539 min
Delta R.T.: 0.002 min
Response: 723243
Conc: 6.34 ng/ml



#29 AR-1254-4


R.T.: 7.643 min
Delta R.T.: 0.008 min
Response: 787369
Conc: 11.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#29 AR-1254-4

R.T.: 6.760 min
Delta R.T.: -0.005 min
Response: 62540
Conc: 1.00 ng/ml




#30 AR-1254-5

R.T.: 8.070 min
Delta R.T.: 0.018 min
Response: 1651344
Conc: 20.09 ng/ml



#30 AR-1254-5


R.T.: 7.181 min
Delta R.T.: -0.006 min
Response: 1434348
Conc: 13.97 ng/ml



#31 AR-1260-1


R.T.: 7.525 min
Delta R.T.: 0.006 min
Response: 407129
Conc: 5.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#31 AR-1260-1

R.T.: 6.664 min
Delta R.T.: -0.004 min
Response: 356757
Conc: 4.36 ng/ml




#32 AR-1260-2

R.T.: 7.773 min
Delta R.T.: 0.000 min
Response: 1433987
Conc: 15.73 ng/ml



#32 AR-1260-2


R.T.: 6.851 min
Delta R.T.: -0.003 min
Response: 543073
Conc: 5.78 ng/ml



#33 AR-1260-3


R.T.: 8.142 min
Delta R.T.: 0.009 min
Response: 284434
Conc: 4.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#33 AR-1260-3

R.T.: 7.008 min
Delta R.T.: -0.005 min
Response: 240053
Conc: 2.65 ng/ml




#34 AR-1260-4

R.T.: 8.378 min
Delta R.T.: 0.007 min
Response: 202028
Conc: 2.68 ng/ml



#34 AR-1260-4


R.T.: 7.484 min
Delta R.T.: -0.004 min
Response: 155645
Conc: 2.33 ng/ml



#35 AR-1260-5


R.T.: 8.710 min
Delta R.T.: 0.002 min
Response: 2121017
Conc: 16.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#35 AR-1260-5

R.T.: 7.724 min
Delta R.T.: -0.002 min
Response: 1134962
Conc: 7.80 ng/ml




#36 AR-1262-1

R.T.: 8.378 min
Delta R.T.: 0.007 min
Response: 202028
Conc: 2.12 ng/ml



#36 AR-1262-1


R.T.: 7.221 min
Delta R.T.: -0.005 min
Response: 177731
Conc: 1.69 ng/ml



#37 AR-1262-2


R.T.: 8.710 min
Delta R.T.: 0.002 min
Response: 2121017
Conc: 13.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#37 AR-1262-2

R.T.: 7.484 min
Delta R.T.: -0.004 min
Response: 155645
Conc: 1.64 ng/ml




#38 AR-1262-3

R.T.: 9.045 min
Delta R.T.: -0.002 min
Response: 421555
Conc: 3.59 ng/ml



#38 AR-1262-3


R.T.: 8.008 min
Delta R.T.: -0.006 min
Response: 147519
Conc: 2.01 ng/ml



#39 AR-1262-4


R.T.: 9.135 min
Delta R.T.: -0.003 min
Response: 461052
Conc: 4.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#39 AR-1262-4

R.T.: 8.074 min
Delta R.T.: -0.004 min
Response: 472163
Conc: 3.58 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.839 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.637 min
Delta R.T.: 0.032 min
Response: 134585
Conc: 2.22 ng/ml



#41 AR-1268-1


R.T.: 9.045 min
Delta R.T.: 0.003 min
Response: 421555
Conc: 2.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#41 AR-1268-1

R.T.: 8.008 min
Delta R.T.: -0.006 min
Response: 147519
Conc: 0.73 ng/ml




#42 AR-1268-2

R.T.: 9.135 min
Delta R.T.: -0.007 min
Response: 461052
Conc: 2.70 ng/ml



#42 AR-1268-2


R.T.: 8.074 min
Delta R.T.: -0.004 min
Response: 472163
Conc: 2.64 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.392 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170




#43 AR-1268-3

R.T.: 8.326 min
Delta R.T.: 0.028 min
Response: 305451
Conc: 1.92 ng/ml




#44 AR-1268-4

R.T.: 9.769 min
Delta R.T.: -0.069 min
Response: 416333
Conc: 6.84 ng/ml



#44 AR-1268-4


R.T.: 8.637 min
Delta R.T.: 0.031 min
Response: 134585
Conc: 2.08 ng/ml



#45 AR-1268-5

R.T.: 10.326 min
Delta R.T.: 0.037 min
Response: 1247772
Conc: 2.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170



#45 AR-1268-5

R.T.: 8.925 min
Delta R.T.: -0.009 min
Response: 742006
Conc: 1.68 ng/ml